

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV052920\
 Data File : VV016330.D
 Acq On : 29 May 2020 23:25
 Operator : SY/MD
 Sample : L2826-03
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 CWKB6

Quant Time: May 30 03:26:28 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR052620WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat May 30 03:05:25 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	87682	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	81889	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	36582	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	25667	4.56	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	91.20%
7) Chloroethane-d5	1.58	69	20172	4.08	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	81.60%
11) 1,1-Dichloroethene-d2	2.13	63	38552	3.65	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	73.00%
20) 2-Butanone-d5	3.94	46	72235	51.71	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.42%
24) Chloroform-d	4.38	84	49285	4.40	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.00%
26) 1,2-Dichloroethane-d4	5.06	65	27722	4.67	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.40%
32) Benzene-d6	5.08	84	98552	4.86	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.20%
36) 1,2-Dichloropropane-d6	6.10	67	30652	4.88	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.60%
41) Toluene-d8	7.34	98	88708	4.59	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.80%
45) trans-1,3-Dichloropropene-	7.65	79	9889	4.23	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.60%
48) 2-Hexanone-d5	8.12	63	53818	50.16	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.32%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	21217	4.57	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.40%
65) 1,2-Dichlorobenzene-d4	11.65	152	30522	5.00	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.00%

Target Compounds

					Ovalue
3) Chloromethane	1.25	50	3547	0.420 ug/L	94
5) Vinyl chloride	1.32	62	2280	0.297 ug/L #	76
12) 1,1-Dichloroethene	2.14	96	3566	0.690 ug/L #	1
14) Carbon disulfide	2.31	76	3167	0.198 ug/L #	94
18) trans-1,2-Dichloroethene	2.78	96	47585	8.408 ug/L	95
19) 1,1-Dichloroethane	3.21	63	2730	0.254 ug/L	95
22) cis-1,2-Dichloroethene	3.94	96	294192	43.951 ug/L	97
34) Trichloroethene	5.94	95	400951	65.312 ug/L	98
49) Tetrachloroethene	8.00	164	1128	0.256 ug/L	93
53) Chlorobenzene	8.90	112	3848	0.242 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV052920\
Data File : VV016330.D
Acq On : 29 May 2020 23:25
Operator : SY/MD
Sample : L2826-03
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
CWKB6

Quant Time: May 30 03:26:28 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR052620WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat May 30 03:05:25 2020
Response via : Initial Calibration

