

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV052920\
 Data File : VV016348.D
 Acq On : 30 May 2020 06:52
 Operator : SY/MD
 Sample : L2788-04
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFXC3

Manual Integrations
 APPROVED

MMDadoda
 6/1/2020 10:10:41 AM

Quant Time: May 30 07:18:30 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR052620WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat May 30 03:05:25 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	74663	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	68473	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	24312	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	21331	4.45	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	89.00%
7) Chloroethane-d5	1.58	69	17412	4.13	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	82.60%
11) 1,1-Dichloroethene-d2	2.12	63	31600	3.51	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	70.20%
20) 2-Butanone-d5	3.97	46	47355m	39.81	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	79.62%
24) Chloroform-d	4.38	84	44210	4.63	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.60%
26) 1,2-Dichloroethane-d4	5.06	65	25355	5.02	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.40%
32) Benzene-d6	5.07	84	87141	5.14	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.80%
36) 1,2-Dichloropropane-d6	6.10	67	27261	5.19	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.80%
41) Toluene-d8	7.34	98	74463	4.61	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.20%
45) trans-1,3-Dichloropropene-	7.65	79	7905	4.04	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.80%
48) 2-Hexanone-d5	8.12	63	43006	47.94	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	95.88%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	16219	4.18	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	83.60%
65) 1,2-Dichlorobenzene-d4	11.65	152	19574	4.82	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV052920\
Data File : VV016348.D
Acq On : 30 May 2020 06:52
Operator : SY/MD
Sample : L2788-04
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFXC3

Manual Integrations
APPROVED

MMDadoda
6/1/2020 10:10:41 AM

Quant Time: May 30 07:18:30 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR052620WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat May 30 03:05:25 2020
Response via : Initial Calibration

