

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV052920\
 Data File : VV016328.D
 Acq On : 29 May 2020 22:29
 Operator : SY/MD
 Sample : VV0529WBL02
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK44

Quant Time: May 30 03:13:37 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR052620WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat May 30 03:05:25 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	93026	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	88883	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	38438	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	28846	4.83	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	96.60%
7) Chloroethane-d5	1.58	69	23388	4.46	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.20%
11) 1,1-Dichloroethene-d2	2.12	63	41356	3.69	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	73.80%
20) 2-Butanone-d5	3.95	46	76669	51.73	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.46%
24) Chloroform-d	4.37	84	55256	4.64	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.80%
26) 1,2-Dichloroethane-d4	5.06	65	30599	4.86	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.20%
32) Benzene-d6	5.07	84	110283	5.01	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.20%
36) 1,2-Dichloropropane-d6	6.09	67	34332	5.04	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.80%
41) Toluene-d8	7.34	98	97819	4.66	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.20%
45) trans-1,3-Dichloropropene-	7.64	79	11498	4.53	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.60%
48) 2-Hexanone-d5	8.12	63	59609	51.19	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	102.38%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	23948	4.76	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.20%
65) 1,2-Dichlorobenzene-d4	11.65	152	34019	5.30	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV052920\
Data File : VV016328.D
Acq On : 29 May 2020 22:29
Operator : SY/MD
Sample : VV0529WBL02
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK44

Quant Time: May 30 03:13:37 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR052620WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat May 30 03:05:25 2020
Response via : Initial Calibration

