

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV052920\
 Data File : VV016342.D
 Acq On : 30 May 2020 04:23
 Operator : SY/MD
 Sample : L2794-02
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFX40

Quant Time: May 30 05:45:22 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR052620WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat May 30 03:05:25 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	76567	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	70567	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	25500	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	23380	4.75	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	95.00%
7) Chloroethane-d5	1.58	69	18312	4.24	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	84.80%
11) 1,1-Dichloroethene-d2	2.12	63	33210	3.60	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.00%
20) 2-Butanone-d5	3.97	46	50288	41.23	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	82.46%
24) Chloroform-d	4.38	84	46863	4.79	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.80%
26) 1,2-Dichloroethane-d4	5.06	65	26493	5.11	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.20%
32) Benzene-d6	5.07	84	92353	5.28	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.60%
36) 1,2-Dichloropropane-d6	6.10	67	29404	5.44	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	108.80%
41) Toluene-d8	7.34	98	79908	4.80	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.00%
45) trans-1,3-Dichloropropene-	7.65	79	8434	4.18	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.60%
48) 2-Hexanone-d5	8.12	63	45816	49.56	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.12%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	17351	4.34	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.80%
65) 1,2-Dichlorobenzene-d4	11.65	152	22144	5.20	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV052920\
Data File : VV016342.D
Acq On : 30 May 2020 04:23
Operator : SY/MD
Sample : L2794-02
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFX40

Quant Time: May 30 05:45:22 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR052620WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat May 30 03:05:25 2020
Response via : Initial Calibration

