

Data File : VV011156.D
Acq On : 31 May 2019 00:45
Operator : SY/MD
Sample : K3081-02
Misc : 4.73G/10ML/MSVOA_V/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GALD2

Quant Time: May 31 01:49:49 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM053019S.M
Quant Title : VOC Analysis
QLast Update : Fri May 31 01:43:46 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	355762	25.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	346623	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	163327	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	97135	20.18	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	80.72%
7) Chloroethane-d5	1.57	69	82560	22.31	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	89.24%
10) 1,1-Dichloroethene-d2	1.97	63	195	0.02	ug/L	-0.15
Spiked Amount	25.000	Range	45 - 110	Recovery	=	0.08%#
20) 2-Butanone-d5	3.94	46	70693	40.18	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	80.36%
24) Chloroform-d	4.40	84	207830	23.90	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	95.60%
26) 1,2-Dichloroethane-d4	5.08	65	125318	22.11	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	88.44%
29) Benzene-d6	5.10	84	420908	22.36	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	89.44%
33) 1,2-Dichloropropane-d6	6.12	67	138553	22.97	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	91.88%
37) Toluene-d8	7.36	98	393709	22.17	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	88.68%
38) trans-1,3-Dichloropropene-	7.66	79	55391	20.29	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	81.16%
39) 2-Hexanone-d5	8.13	63	57208	41.73	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	83.46%
48) 1,1,2,2-Tetrachloroethane-	10.26	84	114968	21.15	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	84.60%
61) 1,2-Dichlorobenzene-d4	11.67	152	141566	23.16	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	92.64%


Target Compounds					Qvalue
13) Acetone	2.19	43	4427	2.163 ug/L	86


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VV011156.D
Acq On : 31 May 2019 00:45
Operator : SY/MD
Sample : K3081-02
Misc : 4.73G/10ML/MSVOA_V/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GALD2

Quant Time: May 31 01:49:49 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM053019S.M
Quant Title : VOC Analysis
QLast Update : Fri May 31 01:43:46 2019
Response via : Initial Calibration





#13

Acetone

Concen: 2.163 ug/L

RT: 2.19 min Scan# 343

Delta R.T. 0.00 min

Lab File: VV011156.D

Acq: 31 May 2019 00:45

Instrument :
MSVOA_V
ClientSampleId :
GALD2

Tgt Ion: 43 Resp: 4427

Ion Ratio Lower Upper

43 100

58 33.9 0.0 53.8

