

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV053020\
 Data File : VV016366.D
 Acq On : 30 May 2020 17:49
 Operator : SY/MD
 Sample : L2826-02DL 10X
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 CWKB5DL

Quant Time: Jun 01 06:03:35 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR053020WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Jun 01 05:08:08 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	93386	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	87265	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	41558	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	31614	3.87	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	77.40%
7) Chloroethane-d5	1.58	69	24888	4.00	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	80.00%
11) 1,1-Dichloroethene-d2	2.13	63	44847	3.14	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	62.80%
20) 2-Butanone-d5	3.94	46	85367	48.39	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	96.78%
24) Chloroform-d	4.38	84	59561	4.31	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.20%
26) 1,2-Dichloroethane-d4	5.06	65	32960	4.58	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.60%
32) Benzene-d6	5.08	84	122377	4.93	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.60%
36) 1,2-Dichloropropane-d6	6.10	67	36927	4.82	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.40%
41) Toluene-d8	7.34	98	107897	4.76	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.20%
45) trans-1,3-Dichloropropene-	7.64	79	12471	4.51	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.20%
48) 2-Hexanone-d5	8.11	63	67966	52.89	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	105.78%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	25149	4.65	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.00%
65) 1,2-Dichlorobenzene-d4	11.65	152	37924	4.91	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.20%

Target Compounds					Ovalue
22) cis-1,2-Dichloroethene	3.94	96	14266	2.026 ug/L	99
34) Trichloroethene	5.94	95	82272	12.145 ug/L	96
49) Tetrachloroethene	8.00	164	2792	0.562 ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV053020\
Data File : VV016366.D
Acq On : 30 May 2020 17:49
Operator : SY/MD
Sample : L2826-02DL 10X
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
CWKB5DL

Quant Time: Jun 01 06:03:35 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR053020WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Jun 01 05:08:08 2020
Response via : Initial Calibration

