

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV053020\
 Data File : VV016367.D
 Acq On : 30 May 2020 18:13
 Operator : SY/MD
 Sample : L2826-03DL 5X
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 CWKB6DL

Quant Time: Jun 01 06:08:49 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR053020WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Jun 01 05:08:08 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	92650	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	85938	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	40480	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	34183	4.21	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	84.20%
7) Chloroethane-d5	1.58	69	26831	4.35	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.00%
11) 1,1-Dichloroethene-d2	2.13	63	48705	3.43	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	68.60%
20) 2-Butanone-d5	3.93	46	86323	49.32	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	98.64%
24) Chloroform-d	4.38	84	59173	4.31	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.20%
26) 1,2-Dichloroethane-d4	5.06	65	32615	4.57	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.40%
32) Benzene-d6	5.08	84	121635	4.98	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.60%
36) 1,2-Dichloropropane-d6	6.10	67	36630	4.86	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.20%
41) Toluene-d8	7.34	98	106066	4.75	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.00%
45) trans-1,3-Dichloropropene-	7.65	79	12325	4.53	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.60%
48) 2-Hexanone-d5	8.11	63	66929	52.89	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	105.78%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	24715	4.64	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.80%
65) 1,2-Dichlorobenzene-d4	11.65	152	36859	4.90	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.00%

Target Compounds

					Ovalue
3) Chloromethane	1.25	50	1392	0.150 ug/L	87
12) 1,1-Dichloroethene	2.13	96	1077	0.181 ug/L	# 1
18) trans-1,2-Dichloroethene	2.78	96	9643	1.524 ug/L	96
22) cis-1,2-Dichloroethene	3.94	96	58629	8.393 ug/L	99
34) Trichloroethene	5.94	95	81652	12.239 ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV053020\
Data File : VV016367.D
Acq On : 30 May 2020 18:13
Operator : SY/MD
Sample : L2826-03DL 5X
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
CWKB6DL

Quant Time: Jun 01 06:08:49 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR053020WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Jun 01 05:08:08 2020
Response via : Initial Calibration

