

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV053020\
 Data File : VV016369.D
 Acq On : 30 May 2020 19:01
 Operator : SY/MD
 Sample : L2826-07DL 10X
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 CWKC0DL

Quant Time: Jun 01 06:19:15 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR053020WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Jun 01 05:08:08 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	84610	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	81377	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	36896	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	34048	4.60	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.00%
7) Chloroethane-d5	1.58	69	26458	4.70	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.00%
11) 1,1-Dichloroethene-d2	2.13	63	47362	3.66	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	73.20%
20) 2-Butanone-d5	3.94	46	84497	52.87	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.74%
24) Chloroform-d	4.38	84	59373	4.74	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.80%
26) 1,2-Dichloroethane-d4	5.06	65	33254	5.10	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.00%
32) Benzene-d6	5.08	84	117873	5.10	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.00%
36) 1,2-Dichloropropane-d6	6.10	67	36060	5.05	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.00%
41) Toluene-d8	7.34	98	105598	4.99	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.80%
45) trans-1,3-Dichloropropene-	7.64	79	12415	4.82	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	96.40%
48) 2-Hexanone-d5	8.11	63	65845	54.95	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	109.90%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	25436	5.04	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.80%
65) 1,2-Dichlorobenzene-d4	11.65	152	36260	5.29	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.80%

Target Compounds					Ovalue
22) cis-1,2-Dichloroethene	3.94	96	14086	2.208 ug/L	90
34) Trichloroethene	5.94	95	50031	7.920 ug/L	94
49) Tetrachloroethene	8.00	164	701	0.151 ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV053020\
Data File : VV016369.D
Acq On : 30 May 2020 19:01
Operator : SY/MD
Sample : L2826-07DL 10X
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
CWKC0DL

Quant Time: Jun 01 06:19:15 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR053020WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Jun 01 05:08:08 2020
Response via : Initial Calibration

