

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV053020\
 Data File : VV016360.D
 Acq On : 30 May 2020 15:25
 Operator : SY/MD
 Sample : VV0530WBL01
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK45

Quant Time: Jun 01 05:26:37 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR053020WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Jun 01 05:08:08 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	83997	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	80860	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	36839	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	35591	4.84	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	96.80%
7) Chloroethane-d5	1.58	69	27468	4.91	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	98.20%
11) 1,1-Dichloroethene-d2	2.12	63	47371	3.68	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	73.60%
20) 2-Butanone-d5	3.93	46	80390	50.67	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	101.34%
24) Chloroform-d	4.37	84	59347	4.77	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.40%
26) 1,2-Dichloroethane-d4	5.06	65	31913	4.93	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.60%
32) Benzene-d6	5.07	84	114785	5.00	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.00%
36) 1,2-Dichloropropane-d6	6.09	67	35418	4.99	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.80%
41) Toluene-d8	7.33	98	99371	4.73	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.60%
45) trans-1,3-Dichloropropene-	7.64	79	12382	4.84	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	96.80%
48) 2-Hexanone-d5	8.11	63	61912	52.00	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	104.00%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	24587	4.90	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.00%
65) 1,2-Dichlorobenzene-d4	11.65	152	34138	4.99	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV053020\
Data File : VV016360.D
Acq On : 30 May 2020 15:25
Operator : SY/MD
Sample : VV0530WBL01
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK45

Quant Time: Jun 01 05:26:37 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR053020WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Jun 01 05:08:08 2020
Response via : Initial Calibration

