

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW053023\
 Data File : VW031276.D
 Acq On : 30 May 2023 18:50
 Operator : SY/MD
 Sample : 02963-02
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0P03

Quant Time: May 31 05:31:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR053023WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed May 31 05:14:27 2023
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene		5.545	114	110999	5.000	ug/L	0.00
28) Chlorobenzene-d5		8.793	117	121906	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4		11.194	152	53544	5.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		1.278	65	45941	4.996	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	100.000%	
7) Chloroethane-d5		1.536	69	52901	5.397	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	108.000%	
11) 1,1-Dichloroethene-d2		2.063	65	23644	4.926	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	98.600%	
20) 2-Butanone-d5		3.815	46	188895	59.469	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	118.940%	
24) Chloroform-d		4.259	84	113549	5.143	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.800%	
26) 1,2-Dichloroethane-d4		4.953	65	67515	5.720	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	114.400%	
32) Benzene-d6		4.970	84	203759	4.622	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.400%	
36) 1,2-Dichloropropane-d6		5.995	67	79952	5.139	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.800%	
41) Toluene-d8		7.252	98	164052	4.178	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.600%	
43) trans-1,3-Dichloroprop...		7.561	79	22765	4.476	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.600%	
46) 2-Hexanone-d5		8.034	63	128191	53.369	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	106.740%	
56) 1,1,2,2-Tetrachloroeth...		10.159	84	61169	5.127	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	102.600%	
66) 1,2-Dichlorobenzene-d4		11.567	152	58377	5.246	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.000%	
Target Compounds							
13) Acetone		2.140	43	4402	1.687	ug/L	91
34) Trichloroethene		5.847	95	15248	1.033	ug/L	94
42) Toluene		7.326	91	55364	0.993	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV053023\
Data File : VV031276.D
Acq On : 30 May 2023 18:50
Operator : SY/MD
Sample : 02963-02
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COP03

Quant Time: May 31 05:31:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR053023WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed May 31 05:14:27 2023
Response via : Initial Calibration

