

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV053023\
 Data File : VV031277.D
 Acq On : 30 May 2023 19:15
 Operator : SY/MD
 Sample : 02963-03
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0P04

Quant Time: May 31 05:31:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR053023WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed May 31 05:14:27 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.545	114	111464	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.793	117	112810	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	48161	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	44318	4.800	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	96.000%
7) Chloroethane-d5	1.536	69	50986	5.180	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	103.600%
11) 1,1-Dichloroethene-d2	2.063	65	23214	4.816	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	96.400%
20) 2-Butanone-d5	3.812	46	173966	54.541	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	109.080%
24) Chloroform-d	4.259	84	111540	5.030	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.600%
26) 1,2-Dichloroethane-d4	4.950	65	64394	5.433	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	108.600%
32) Benzene-d6	4.970	84	197342	4.837	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.800%
36) 1,2-Dichloropropane-d6	5.998	67	78010	5.418	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	108.400%
41) Toluene-d8	7.252	98	157511	4.335	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.800%
43) trans-1,3-Dichloroprop...	7.564	79	23882	5.075	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	101.400%
46) 2-Hexanone-d5	8.034	63	120979	54.428	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	108.860%
56) 1,1,2,2-Tetrachloroeth...	10.159	84	59917	5.427	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	108.600%
66) 1,2-Dichlorobenzene-d4	11.567	152	57355	5.731	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	114.600%
Target Compounds						
					Qvalue	
13) Acetone	2.137	43	4633	1.768	ug/L	86
34) Trichloroethene	5.841	95	20026	1.466	ug/L	97
42) Toluene	7.323	91	134303	2.604	ug/L	99
53) m,p-Xylene	9.082	106	2025	0.104	ug/L	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV053023\
Data File : VV031277.D
Acq On : 30 May 2023 19:15
Operator : SY/MD
Sample : 02963-03
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0P04

Quant Time: May 31 05:31:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR053023WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed May 31 05:14:27 2023
Response via : Initial Calibration

