

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW053124\
 Data File : VW035938.D
 Acq On : 31 May 2024 18:21
 Operator : SY/MD
 Sample : P2663-09
 Misc : 5.71g/10.0mL/MSVOA_V/SOIL/A
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BH696

Quant Time: May 31 23:05:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM053024SMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri May 31 22:59:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	490556	25.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	538628	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.182	152	238094	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	189886	21.054	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	84.200%		
7) Chloroethane-d5	1.529	69	169294	23.541	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	94.160%		
11) 1,1-Dichloroethene-d2	2.053	65	86070	21.437	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	85.760%		
21) 2-Butanone-d5	3.838	46	84886	41.022	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	82.040%		
24) Chloroform-d	4.246	84	399610	25.623	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	102.480%		
26) 1,2-Dichloroethane-d4	4.941	65	226765	27.732	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	110.920%		
32) Benzene-d6	4.957	84	736768	23.012	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	92.040%		
36) 1,2-Dichloropropane-d6	5.986	67	256754	25.465	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	101.880%		
41) Toluene-d8	7.240	98	595954	21.096	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	84.400%		
43) trans-1,3-Dichloroprop...	7.551	79	92606	22.401	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	89.600%		
47) 2-Hexanone-d5	8.031	63	68718	50.304	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	100.600%		
56) 1,1,2,2-Tetrachloroeth...	10.149	84	212672	28.227	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	112.920%		
66) 1,2-Dichlorobenzene-d4	11.558	152	240233	27.514	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	110.040%		
Target Compounds						Qvalue
16) Methylene chloride	2.442	84	89534	8.066	ug/L	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV053124\
Data File : VV035938.D
Acq On : 31 May 2024 18:21
Operator : SY/MD
Sample : P2663-09
Misc : 5.71g/10.0mL/MSVOA_V/SOIL/A
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BH696

Quant Time: May 31 23:05:16 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM053024SMA.M
Quant Title : VOC Analysis
QLast Update : Fri May 31 22:59:48 2024
Response via : Initial Calibration





#16
Methylene chloride
Concen: 8.066 ug/L
RT: 2.442 min Scan# 41
Delta R.T. 0.000 min
Lab File: VV035938.D
Acq: 31 May 2024 18:21

Instrument :
MSVOA_V
ClientSampleId :
BH696

Tgt Ion: 84 Resp: 89534
Ion Ratio Lower Upper
84 100
86 63.9 44.6 82.8
49 116.0 81.1 150.5

