

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV060121\
 Data File : VV021461.D
 Acq On : 01 Jun 2021 15:36
 Operator : SY/MD
 Sample : M2568-07 50X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0GN4

Quant Time: Jun 02 00:50:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR052521WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jun 02 00:49:39 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	230788	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	213381	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	121739	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	40051	4.102	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	82.000%
7) Chloroethane-d5	1.564	69	44093	4.732	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.600%
11) 1,1-Dichloroethene-d2	2.101	63	59724	3.227	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	64.600%
20) 2-Butanone-d5	3.924	46	203509	57.982	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	115.960%
24) Chloroform-d	4.349	84	135480	5.003	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.000%
26) 1,2-Dichloroethane-d4	5.037	65	65927	5.295	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.000%
32) Benzene-d6	5.050	84	257099	5.008	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.200%
36) 1,2-Dichloropropane-d6	6.072	67	83283	5.002	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.000%
41) Toluene-d8	7.317	98	229319	4.871	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.400%
43) trans-1,3-Dichloroprop...	7.628	79	33781	4.875	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	97.400%
46) 2-Hexanone-d5	8.091	63	178751	54.782	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	109.560%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	69735	5.197	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	104.000%
66) 1,2-Dichlorobenzene-d4	11.625	152	109549	5.221	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.400%
Target Compounds						Qvalue
16) Methylene chloride	2.507	84	16841	0.895	ug/L	97
18) trans-1,2-Dichloroethene	2.761	96	3977	0.278	ug/L	89
22) cis-1,2-Dichloroethene	3.915	96	18189	1.170	ug/L	92
31) Carbon tetrachloride	4.828	117	2382	0.110	ug/L	95
34) Trichloroethene	5.915	95	5831143	381.262	ug/L	99
45) 1,1,2-Trichloroethane	7.850	97	2943	0.288	ug/L	90
47) Tetrachloroethene	7.979	164	9291	0.666	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV060121\
Data File : VV021461.D
Acq On : 01 Jun 2021 15:36
Operator : SY/MD
Sample : M2568-07 50X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0GN4

Quant Time: Jun 02 00:50:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR052521WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jun 02 00:49:39 2021
Response via : Initial Calibration

