

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV060220\
 Data File : VV016425.D
 Acq On : 02 Jun 2020 12:32
 Operator : SY/MD
 Sample : L2780-07
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Jun 02 18:31:02 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR053020WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jun 02 18:21:27 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	83385	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	79841	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	35280	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	28104	3.85	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	77.00%
7) Chloroethane-d5	1.58	69	23790	4.29	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	85.80%
11) 1,1-Dichloroethene-d2	2.12	63	41431	3.25	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.00%
20) 2-Butanone-d5	3.94	46	85489	54.28	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	108.56%
24) Chloroform-d	4.37	84	53965	4.37	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.40%
26) 1,2-Dichloroethane-d4	5.05	65	32211	5.01	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.20%
32) Benzene-d6	5.07	84	111847	4.93	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.60%
36) 1,2-Dichloropropane-d6	6.09	67	34803	4.97	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.40%
41) Toluene-d8	7.33	98	92974	4.48	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.60%
45) trans-1,3-Dichloropropene-	7.64	79	11316	4.48	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.60%
48) 2-Hexanone-d5	8.11	63	63522	54.03	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	108.06%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	24801	5.01	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.20%
65) 1,2-Dichlorobenzene-d4	11.65	152	32578	4.97	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV060220\
Data File : VV016425.D
Acq On : 02 Jun 2020 12:32
Operator : SY/MD
Sample : L2780-07
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Jun 02 18:31:02 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR053020WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jun 02 18:21:27 2020
Response via : Initial Calibration

