

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV060221\
 Data File : VV021470.D
 Acq On : 02 Jun 2021 10:32
 Operator : SY/MD
 Sample : VV0602WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK183

Quant Time: Jun 03 02:16:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR052521WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jun 03 02:14:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	229575	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	211609	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	118499	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	57466	5.916	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 118.400%		
7) Chloroethane-d5	1.564	69	46842	5.054	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 101.000%		
11) 1,1-Dichloroethene-d2	2.104	63	73926	4.015	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 80.400%		
20) 2-Butanone-d5	3.931	46	133609	38.268	ug/L	0.01
Spiked Amount	50.000	Range 40 - 130	Recovery	= 76.540%		
24) Chloroform-d	4.352	84	117588	4.365	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 87.400%		
26) 1,2-Dichloroethane-d4	5.037	65	56554	4.566	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 91.400%		
32) Benzene-d6	5.053	84	248627	4.884	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 97.600%		
36) 1,2-Dichloropropane-d6	6.072	67	74184	4.493	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 89.800%		
41) Toluene-d8	7.316	98	238476	5.108	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 102.200%		
43) trans-1,3-Dichloroprop...	7.628	79	30080	4.377	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 87.600%		
46) 2-Hexanone-d5	8.095	63	118715	36.687	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 73.380%		
56) 1,1,2,2-Tetrachloroeth...	10.217	84	54459	4.092	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 81.800%		
66) 1,2-Dichlorobenzene-d4	11.625	152	97859	4.791	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 95.800%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV060221\
Data File : VV021470.D
Acq On : 02 Jun 2021 10:32
Operator : SY/MD
Sample : VV0602WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK183

Quant Time: Jun 03 02:16:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR052521WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jun 03 02:14:36 2021
Response via : Initial Calibration

