

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV060221\
 Data File : VV021473.D
 Acq On : 02 Jun 2021 11:56
 Operator : SY/MD
 Sample : M2568-01DL 10X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0GN0DL

Quant Time: Jun 03 02:16:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR052521WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jun 03 02:14:36 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	220682	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	208675	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	118097	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	56121	6.011	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 120.200%		
7) Chloroethane-d5	1.565	69	45844	5.146	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 103.000%		
11) 1,1-Dichloroethene-d2	2.105	63	72384	4.090	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 81.800%		
20) 2-Butanone-d5	3.934	46	113724	33.885	ug/L	0.02
Spiked Amount	50.000	Range 40 - 130	Recovery	= 67.760%		
24) Chloroform-d	4.352	84	115081	4.444	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 88.800%		
26) 1,2-Dichloroethane-d4	5.040	65	54305	4.562	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 91.200%		
32) Benzene-d6	5.053	84	243141	4.843	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 96.800%		
36) 1,2-Dichloropropane-d6	6.072	67	72664	4.463	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 89.200%		
41) Toluene-d8	7.320	98	233145	5.064	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 101.200%		
43) trans-1,3-Dichloroprop...	7.629	79	29073	4.290	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 85.800%		
46) 2-Hexanone-d5	8.095	63	116020	36.359	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 72.720%		
56) 1,1,2,2-Tetrachloroeth...	10.220	84	53806	4.100	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 82.000%		
66) 1,2-Dichlorobenzene-d4	11.625	152	96590	4.745	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 95.000%		
Target Compounds						Qvalue
22) cis-1,2-Dichloroethene	3.921	96	3148	0.212	ug/L	92
34) Trichloroethene	5.915	95	116344	7.779	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV060221\
Data File : VV021473.D
Acq On : 02 Jun 2021 11:56
Operator : SY/MD
Sample : M2568-01DL 10X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0GN0DL

Quant Time: Jun 03 02:16:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR052521WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jun 03 02:14:36 2021
Response via : Initial Calibration

