




Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV060320\
Data File : VV016489.D
Acq On : 04 Jun 2020 04:48
Operator : SY/MD
Sample : L2849-11
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 68 Sample Multiplier: 1

Quant Time: Jun 04 05:39:04 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060320WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jun 04 05:34:42 2020
Response via : Initial Calibration



TIC: VV016489.D

(12) 1,1-Dichloroethene (T)

1.843min (-0.289) 0.02ug/L

response 40

Ion	Exp%	Act%
96.00	100	100
61.00	179.30	148.57
63.00	148.50	119.05
0.00	0.00	0.00