

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV060320\
 Data File : VV016488.D
 Acq On : 04 Jun 2020 04:25
 Operator : SY/MD
 Sample : L2849-10
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 67 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ETTR0

Quant Time: Jun 04 07:26:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060320WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jun 04 05:34:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	364148	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	345216	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	153298	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	117177	45.43	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	90.86%
7) Chloroethane-d5	1.57	69	81300	43.66	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	87.32%
11) 1,1-Dichloroethene-d2	2.12	63	153491	33.57	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	67.14%
21) 2-Butanone-d5	3.90	46	176409	106.97	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	106.97%
24) Chloroform-d	4.37	84	225675	46.62	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.24%
26) 1,2-Dichloroethane-d4	5.06	65	158546	48.92	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.84%
32) Benzene-d6	5.07	84	452687	48.95	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.90%
36) 1,2-Dichloropropane-d6	6.09	67	139194	48.31	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.62%
41) Toluene-d8	7.34	98	401475	47.80	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.60%
43) trans-1,3-Dichloropropene-	7.64	79	58012	41.31	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.62%
47) 2-Hexanone-d5	8.11	63	131867	106.86	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	106.86%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	185960	47.59	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.18%
64) 1,2-Dichlorobenzene-d4	11.65	152	155992	54.15	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	108.30%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV060320\
Data File : VV016488.D
Acq On : 04 Jun 2020 04:25
Operator : SY/MD
Sample : L2849-10
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 67 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ETTR0

Quant Time: Jun 04 07:26:27 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060320WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jun 04 05:34:42 2020
Response via : Initial Calibration

