

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV060320\
 Data File : VV016492.D
 Acq On : 04 Jun 2020 05:59
 Operator : SY/MD
 Sample : VIBLK51
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 71 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK51

Quant Time: Jun 04 07:55:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060320WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jun 04 05:34:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	378024	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	357292	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	167989	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	103076	38.50	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	77.00%
7) Chloroethane-d5	1.58	69	83997	43.46	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	86.92%
11) 1,1-Dichloroethene-d2	2.12	63	152028	32.03	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	64.06%
21) 2-Butanone-d5	3.89	46	176323	102.99	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	102.99%
24) Chloroform-d	4.38	84	231215	46.02	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.04%
26) 1,2-Dichloroethane-d4	5.06	65	161569	48.02	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.04%
32) Benzene-d6	5.07	84	462751	48.35	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.70%
36) 1,2-Dichloropropane-d6	6.09	67	145230	48.70	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.40%
41) Toluene-d8	7.34	98	409310	47.09	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.18%
43) trans-1,3-Dichloropropene-	7.64	79	61889	42.58	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.16%
47) 2-Hexanone-d5	8.11	63	139834	109.49	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	109.49%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	193144	47.75	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.50%
64) 1,2-Dichlorobenzene-d4	11.65	152	167120	52.94	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.88%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV060320\
Data File : VV016492.D
Acq On : 04 Jun 2020 05:59
Operator : SY/MD
Sample : VIBLK51
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 71 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK51

Quant Time: Jun 04 07:55:31 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060320WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jun 04 05:34:42 2020
Response via : Initial Calibration

