

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV060421\
 Data File : VV021503.D
 Acq On : 04 Jun 2021 13:44
 Operator : SY/MD
 Sample : M2605-09
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG7H9

Quant Time: Jun 05 03:58:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM052721WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Jun 05 03:34:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	466591	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	438474	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	248181	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	110400	35.938	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	71.880%
7) Chloroethane-d5	1.568	69	89064	38.975	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	77.940%
11) 1,1-Dichloroethene-d2	2.108	63	155198	29.566	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	59.140%#
21) 2-Butanone-d5	3.886	46	218400	99.144	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	99.140%
24) Chloroform-d	4.352	84	261244	42.720	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.440%
26) 1,2-Dichloroethane-d4	5.034	65	161683	45.903	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.800%
32) Benzene-d6	5.053	84	526211	42.899	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.800%
36) 1,2-Dichloropropane-d6	6.072	67	168794	44.559	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	89.120%
41) Toluene-d8	7.316	98	498808	42.683	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.360%
43) trans-1,3-Dichloroprop...	7.625	79	86235	42.539	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.080%
47) 2-Hexanone-d5	8.088	63	156931	90.221	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	90.220%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	229109	43.571	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	87.140%
66) 1,2-Dichlorobenzene-d4	11.625	152	229551	44.495	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.000%
Target Compounds						
					Qvalue	
13) Acetone	2.178	43	12361	6.522	ug/L #	84
46) Tetrachloroethene	7.982	164	2495	0.796	ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV060421\
Data File : VV021503.D
Acq On : 04 Jun 2021 13:44
Operator : SY/MD
Sample : M2605-09
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG7H9

Quant Time: Jun 05 03:58:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM052721WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jun 05 03:34:55 2021
Response via : Initial Calibration

