

Data File : VV016526.D
Acq On : 05 Jun 2020 11:23
Operator : SY/MD
Sample : L2836-06
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Jun 06 00:35:13 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060320WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jun 06 00:07:58 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	373346	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	353526	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	163802	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	115508	43.68	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	87.36%		
7) Chloroethane-d5	1.56	69	84162	44.09	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	88.18%		
11) 1,1-Dichloroethene-d2	2.12	63	160318	34.20	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	68.40%		
21) 2-Butanone-d5	3.89	46	195347	115.53	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	115.53%		
24) Chloroform-d	4.37	84	247223	49.82	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	99.64%		
26) 1,2-Dichloroethane-d4	5.05	65	173578	52.23	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	104.46%		
32) Benzene-d6	5.07	84	492106	51.96	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	103.92%		
36) 1,2-Dichloropropane-d6	6.09	67	155912	52.84	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	105.68%		
41) Toluene-d8	7.34	98	439351	51.08	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	102.16%		
43) trans-1,3-Dichloropropene-	7.64	79	71258	49.55	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	99.10%		
47) 2-Hexanone-d5	8.11	63	145696	115.29	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	115.29%		
57) 1,1,2,2-Tetrachloroethane-	10.24	84	207301	51.80	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	103.60%		
64) 1,2-Dichlorobenzene-d4	11.65	152	178626	58.03	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	116.06%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VV016526.D
Acq On : 05 Jun 2020 11:23
Operator : SY/MD
Sample : L2836-06
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Jun 06 00:35:13 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060320WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jun 06 00:07:58 2020
Response via : Initial Calibration

