

Data File : VV016527.D
Acq On : 05 Jun 2020 11:45
Operator : SY/MD
Sample : L2844-12
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK02

Quant Time: Jun 06 00:41:05 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060320WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jun 06 00:07:58 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	376939	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	356014	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	170327	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	110584	41.42	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	82.84%
7) Chloroethane-d5	1.57	69	81814	42.45	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	84.90%
11) 1,1-Dichloroethene-d2	2.12	63	155617	32.88	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	65.76%
21) 2-Butanone-d5	3.90	46	181986	106.60	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	106.60%
24) Chloroform-d	4.38	84	238312	47.56	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.12%
26) 1,2-Dichloroethane-d4	5.06	65	165472	49.32	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.64%
32) Benzene-d6	5.07	84	473174	49.62	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.24%
36) 1,2-Dichloropropane-d6	6.09	67	148756	50.06	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.12%
41) Toluene-d8	7.34	98	426089	49.20	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.40%
43) trans-1,3-Dichloropropene-	7.64	79	69675	48.11	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.22%
47) 2-Hexanone-d5	8.11	63	138299	108.67	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	108.67%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	200550	49.76	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	99.52%
64) 1,2-Dichlorobenzene-d4	11.65	152	173992	54.36	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	108.72%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VV016527.D
Acq On : 05 Jun 2020 11:45
Operator : SY/MD
Sample : L2844-12
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK02

Quant Time: Jun 06 00:41:05 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060320WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jun 06 00:07:58 2020
Response via : Initial Calibration

