

Data File : VV016533.D
Acq On : 05 Jun 2020 14:08
Operator : SY/MD
Sample : L2861-03
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F9D77

Quant Time: Jun 06 01:09:24 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060320WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jun 06 00:07:58 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	348022	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	328199	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	154703	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	111081	45.06	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	90.12%
7) Chloroethane-d5	1.57	69	80531	45.25	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	90.50%
11) 1,1-Dichloroethene-d2	2.12	63	147945	33.86	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	67.72%
21) 2-Butanone-d5	3.89	46	170496	108.17	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	108.17%
24) Chloroform-d	4.37	84	221094	47.79	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.58%
26) 1,2-Dichloroethane-d4	5.05	65	154517	49.88	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.76%
32) Benzene-d6	5.07	84	443859	50.49	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.98%
36) 1,2-Dichloropropane-d6	6.09	67	138729	50.64	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.28%
41) Toluene-d8	7.34	98	387964	48.59	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.18%
43) trans-1,3-Dichloropropene-	7.64	79	59472	44.55	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.10%
47) 2-Hexanone-d5	8.11	63	124290	105.94	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	105.94%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	182095	49.01	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	98.02%
64) 1,2-Dichlorobenzene-d4	11.65	152	158348	54.47	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	108.94%

Target Compounds					Qvalue
13) Acetone	2.17	43	2590	1.758 ug/L	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VV016533.D
Acq On : 05 Jun 2020 14:08
Operator : SY/MD
Sample : L2861-03
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F9D77

Quant Time: Jun 06 01:09:24 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060320WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jun 06 00:07:58 2020
Response via : Initial Calibration

