

Data File : VV016570.D
Acq On : 06 Jun 2020 05:38
Operator : SY/MD
Sample : L2862-16
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F9E32

Quant Time: Jun 06 07:45:58 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060320WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jun 06 06:28:49 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	368357	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	347133	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	165823	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	112939	43.29	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	86.58%
7) Chloroethane-d5	1.56	69	84542	44.89	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	89.78%
11) 1,1-Dichloroethene-d2	2.12	63	156792	33.90	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	67.80%
21) 2-Butanone-d5	3.89	46	201075	120.53	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	120.53%
24) Chloroform-d	4.37	84	244592	49.96	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.92%
26) 1,2-Dichloroethane-d4	5.05	65	171404	52.28	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.56%
32) Benzene-d6	5.07	84	493135	53.03	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.06%
36) 1,2-Dichloropropane-d6	6.09	67	155816	53.78	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	107.56%
41) Toluene-d8	7.34	98	438732	51.95	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.90%
43) trans-1,3-Dichloropropene-	7.64	79	70424	49.87	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	99.74%
47) 2-Hexanone-d5	8.11	63	152350	122.78	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	122.78%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	209287	53.26	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	106.52%
64) 1,2-Dichlorobenzene-d4	11.65	152	182142	58.45	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	116.90%

Target Compounds					Qvalue
13) Acetone	2.17	43	6334	4.062 ug/L	95
27) 1,2-Dichloroethane	5.16	62	19461	4.788 ug/L	98
33) Benzene	5.13	78	50462	4.943 ug/L	100
42) Toluene	7.41	91	7904	0.731 ug/L	92
52) Ethylbenzene	9.04	91	12306	1.045 ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VV016570.D
Acq On : 06 Jun 2020 05:38
Operator : SY/MD
Sample : L2862-16
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F9E32

Quant Time: Jun 06 07:45:58 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060320WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jun 06 06:28:49 2020
Response via : Initial Calibration

