

Data File : VV016525.D
Acq On : 05 Jun 2020 10:59
Operator : SY/MD
Sample : VV0605WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK68

Quant Time: Jun 06 00:30:58 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060320WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jun 06 00:07:58 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	398990	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	375304	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	177896	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	119377	42.24	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	84.48%
7) Chloroethane-d5	1.58	69	89400	43.82	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	87.64%
11) 1,1-Dichloroethene-d2	2.12	63	164183	32.77	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	65.54%
21) 2-Butanone-d5	3.89	46	191318	105.88	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	105.88%
24) Chloroform-d	4.37	84	251040	47.34	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.68%
26) 1,2-Dichloroethane-d4	5.05	65	173236	48.78	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.56%
32) Benzene-d6	5.07	84	503099	50.04	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.08%
36) 1,2-Dichloropropane-d6	6.09	67	157200	50.18	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.36%
41) Toluene-d8	7.33	98	452452	49.55	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.10%
43) trans-1,3-Dichloropropene-	7.64	79	74075	48.52	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	97.04%
47) 2-Hexanone-d5	8.10	63	144257	107.53	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	107.53%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	205389	48.34	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	96.68%
64) 1,2-Dichlorobenzene-d4	11.65	152	177730	53.16	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	106.32%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VV016525.D
Acq On : 05 Jun 2020 10:59
Operator : SY/MD
Sample : VV0605WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK68

Quant Time: Jun 06 00:30:58 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060320WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jun 06 00:07:58 2020
Response via : Initial Calibration

