

Data File : VV016528.D
Acq On : 05 Jun 2020 12:09
Operator : SY/MD
Sample : L2868-02
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Jun 06 00:44:20 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060320WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jun 06 00:07:58 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	363507	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	346084	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	165303	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	110791	43.03	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	86.06%
7) Chloroethane-d5	1.56	69	80032	43.06	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	86.12%
11) 1,1-Dichloroethene-d2	2.12	63	152738	33.47	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	66.94%
21) 2-Butanone-d5	3.89	46	178919	108.68	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	108.68%
24) Chloroform-d	4.37	84	231440	47.90	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.80%
26) 1,2-Dichloroethane-d4	5.05	65	164763	50.92	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.84%
32) Benzene-d6	5.07	84	464124	50.06	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.12%
36) 1,2-Dichloropropane-d6	6.09	67	143750	49.76	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.52%
41) Toluene-d8	7.34	98	418850	49.75	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.50%
43) trans-1,3-Dichloropropene-	7.64	79	66445	47.20	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.40%
47) 2-Hexanone-d5	8.11	63	137939	111.50	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	111.50%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	196216	50.08	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	100.16%
64) 1,2-Dichlorobenzene-d4	11.65	152	170821	54.99	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	109.98%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VV016528.D
Acq On : 05 Jun 2020 12:09
Operator : SY/MD
Sample : L2868-02
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Jun 06 00:44:20 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060320WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jun 06 00:07:58 2020
Response via : Initial Calibration

