

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW060523\
 Data File : VW031358.D
 Acq On : 05 Jun 2023 19:41
 Operator : SY/MD
 Sample : 03000-06MS 20X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F6C29MS

Quant Time: Jun 06 02:11:46 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR060123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Jun 06 02:05:53 2023
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Krupa Patel 06/06/2023
 Supervised By :Mahesh Dadoda 06/08/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.539	114	138882	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.793	117	129583	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.191	152	75681	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	49207	3.483	ug/L	0.00
Spiked Amount 5.000	Range 40 - 130		Recovery =	69.600%		
7) Chloroethane-d5	1.529	69	44088	3.695	ug/L	0.00
Spiked Amount 5.000	Range 65 - 130		Recovery =	73.800%		
11) 1,1-Dichloroethene-d2	2.053	65	26361	3.843	ug/L	0.00
Spiked Amount 5.000	Range 60 - 125		Recovery =	76.800%		
20) 2-Butanone-d5	3.864	46	151790	46.421	ug/L	0.03
Spiked Amount 50.000	Range 40 - 130		Recovery =	92.840%		
24) Chloroform-d	4.256	84	108590	4.572	ug/L	0.00
Spiked Amount 5.000	Range 70 - 125		Recovery =	91.400%		
26) 1,2-Dichloroethane-d4	4.950	65	57831	4.540	ug/L	0.00
Spiked Amount 5.000	Range 70 - 130		Recovery =	90.800%		
32) Benzene-d6	4.963	84	192666	4.891	ug/L	0.00
Spiked Amount 5.000	Range 70 - 125		Recovery =	97.800%		
36) 1,2-Dichloropropane-d6	5.995	67	63562	4.777	ug/L	0.00
Spiked Amount 5.000	Range 60 - 140		Recovery =	95.600%		
41) Toluene-d8	7.249	98	174346	4.713	ug/L	0.00
Spiked Amount 5.000	Range 70 - 130		Recovery =	94.200%		
43) trans-1,3-Dichloroprop...	7.561	79	23122	4.773	ug/L	0.00
Spiked Amount 5.000	Range 55 - 130		Recovery =	95.400%		
46) 2-Hexanone-d5	8.037	63	106808	54.886	ug/L	0.00
Spiked Amount 50.000	Range 45 - 130		Recovery =	109.780%		
56) 1,1,2,2-Tetrachloroeth...	10.159	84	49535	4.897	ug/L	0.00
Spiked Amount 5.000	Range 65 - 120		Recovery =	98.000%		
66) 1,2-Dichlorobenzene-d4	11.567	152	64857	4.544	ug/L	0.00
Spiked Amount 5.000	Range 80 - 120		Recovery =	90.800%		
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.105	85	78756	4.202	ug/L	99
3) Chloromethane	1.214	50	71354	3.853	ug/L	97
5) Vinyl chloride	1.278	62	83588	4.516	ug/L	99
6) Bromomethane	1.484	94	41840	3.943	ug/L	96
8) Chloroethane	1.545	64	45443	3.937	ug/L	98
9) Trichlorofluoromethane	1.709	101	120689	4.427	ug/L	99
10) 1,1,2-Trichloro-1,2,2-...	2.063	101	62602	4.154	ug/L	97
12) 1,1-Dichloroethene	2.066	96	55829	4.303	ug/L	92
13) Acetone	2.224	43	90653	23.990	ug/L	56
14) Carbon disulfide	2.237	76	158197	4.113	ug/L	100
15) Methyl Acetate	2.413	43	26037m	3.547	ug/L	
16) Methylene chloride	2.446	84	66642	4.225	ug/L	98
17) Methyl tert-butyl Ether	2.716	73	140275	4.744	ug/L #	52
18) trans-1,2-Dichloroethene	2.693	96	57342	4.660	ug/L	96
19) 1,1-Dichloroethane	3.114	63	116515	4.263	ug/L	100
21) 2-Butanone	3.944	43	171807m	34.960	ug/L	
22) cis-1,2-Dichloroethene	3.818	96	270322	21.680	ug/L	99
23) Bromochloromethane	4.153	128	26271	4.457	ug/L	92
25) Chloroform	4.282	83	116194	4.066	ug/L	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW060523\
 Data File : VW031358.D
 Acq On : 05 Jun 2023 19:41
 Operator : SY/MD
 Sample : 03000-06MS 20X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F6C29MS


Quant Time: Jun 06 02:11:46 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR060123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Jun 06 02:05:53 2023
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 06/06/2023
 Supervised By :Mahesh Dadoda 06/08/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
27) 1,2-Dichloroethane	5.050	62	83451	5.024	ug/L	99
29) 1,1,1-Trichloroethane	4.516	97	106198	4.929	ug/L	99
30) Cyclohexane	4.584	56	593677	30.349	ug/L	96
31) Carbon tetrachloride	4.738	117	94352	4.910	ug/L	98
33) Benzene	5.011	78	978193	21.211	ug/L	100
34) Trichloroethene	5.835	95	413641	33.210	ug/L	97
35) Methylcyclohexane	6.053	83	317236	17.109	ug/L	97
37) 1,2-Dichloropropane	6.098	63	63286	4.663	ug/L	98
38) Bromodichloromethane	6.439	83	81429	4.692	ug/L	99
39) cis-1,3-Dichloropropene	6.963	75	87577	4.916	ug/L	97
40) 4-Methyl-2-pentanone	7.169	43	428712	53.835	ug/L	96
42) Toluene	7.320	91	253191	5.174	ug/L	98
44) trans-1,3-Dichloropropene	7.587	75	76304	4.979	ug/L	99
45) 1,1,2-Trichloroethane	7.777	97	43153	4.684	ug/L	98
47) Tetrachloroethene	7.908	164	109668	11.299	ug/L	99
48) 2-Hexanone	8.088	43	307285	48.587	ug/L	99
49) Dibromochloromethane	8.182	129	50646	4.923	ug/L	98
50) 1,2-Dibromoethane	8.291	107	42123	4.944	ug/L	95
51) Chlorobenzene	8.822	112	147954	4.854	ug/L	97
52) Ethylbenzene	8.953	91	453607	8.862	ug/L	98
53) m,p-Xylene	9.079	106	326982	17.756	ug/L	98
54) o-Xylene	9.484	106	125199	7.254	ug/L	99
55) Styrene	9.503	104	169108	5.509	ug/L	97
57) 1,1,2,2-Tetrachloroethane	10.185	83	52976	4.824	ug/L	99
59) Bromoform	9.674	173	29753	4.879	ug/L	97
60) Isopropylbenzene	9.873	105	284614	5.235	ug/L	99
61) 1,2,3-Trichloropropane	10.217	75	39693	4.293	ug/L	97
62) 1,3,5-Trimethylbenzene	10.484	105	269746	6.123	ug/L	98
63) 1,2,4-Trimethylbenzene	10.860	105	338226	7.438	ug/L	96
64) 1,3-Dichlorobenzene	11.124	146	128923	4.897	ug/L	96
65) 1,4-Dichlorobenzene	11.217	146	129308	4.689	ug/L	99
67) 1,2-Dichlorobenzene	11.587	146	120524	4.772	ug/L	99
68) 1,2-Dibromo-3-chloropr...	12.374	75	9241	4.357	ug/L	90
69) 1,3,5-Trichlorobenzene	12.590	180	93492	4.779	ug/L	98
70) 1,2,4-trichlorobenzene	13.207	180	74416	4.743	ug/L	98
71) Naphthalene	13.445	128	123412	5.127	ug/L	98
72) 1,2,3-Trichlorobenzene	13.689	180	68024	4.833	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed







#2
 Dichlorodifluoromethane
 Concen: 4.202 ug/L
 RT: 1.105 min Scan# 20
 Delta R.T. 0.000 min
 Lab File: VV031358.D
 Acq: 05 Jun 2023 19:41

Instrument :
 MSVOA_V
 ClientSampleId :
 F6C29MS

Tgt Ion: 85 Resp: 78750
 Ion Ratio Lower Upper
 85 100
 87 32.5 25.6 38.4




Manual Integrations
 APPROVED


Reviewed By :Krupa Patel 06/06/2023
 Supervised By :Mahesh Dadoda 06/08/2023



#3
 Chloromethane
 Concen: 3.853 ug/L
 RT: 1.214 min Scan# 54
 Delta R.T. -0.000 min
 Lab File: VV031358.D
 Acq: 05 Jun 2023 19:41

Tgt Ion: 50 Resp: 71354
 Ion Ratio Lower Upper
 50 100
 52 31.6 23.2 43.2









#5
 Vinyl chloride
 Concen: 4.516 ug/L
 RT: 1.278 min Scan# 74
 Delta R.T. -0.003 min
 Lab File: VV031358.D
 Acq: 05 Jun 2023 19:41

Instrument :
 MSVOA_V
 ClientSampleId :
 F6C29MS

Tgt Ion: 62 Resp: 8358
 Ion Ratio Lower Upper
 62 100
 64 31.8 22.0 40.8




Manual Integrations
 APPROVED


Reviewed By :Krupa Patel 06/06/2023
 Supervised By :Mahesh Dadoda 06/08/2023



#6
 Bromomethane
 Concen: 3.943 ug/L
 RT: 1.484 min Scan# 138
 Delta R.T. -0.003 min
 Lab File: VV031358.D
 Acq: 05 Jun 2023 19:41

Tgt Ion: 94 Resp: 41840
 Ion Ratio Lower Upper
 94 100
 96 94.9 63.7 118.3









#8
Chloroethane
Concen: 3.937 ug/L
RT: 1.545 min Scan# 11
Delta R.T. -0.003 min
Lab File: VV031358.D
Acq: 05 Jun 2023 19:41

Instrument :
MSVOA_V
ClientSampleId :
F6C29MS

Tgt Ion: 64 Resp: 4544
Ion Ratio Lower Upper
64 100
66 33.4 22.5 41.9




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/06/2023
Supervised By :Mahesh Dadoda 06/08/2023



#9
Trichlorofluoromethane
Concen: 4.427 ug/L
RT: 1.709 min Scan# 208
Delta R.T. -0.003 min
Lab File: VV031358.D
Acq: 05 Jun 2023 19:41

Tgt Ion:101 Resp: 120689
Ion Ratio Lower Upper
101 100
103 65.1 51.4 77.0





#10

1,1,2-Trichloro-1,2,2-trifluoroethane

Concen: 4.154 ug/L

RT: 2.063 min Scan# 318

Delta R.T. -0.006 min

Lab File: VV031358.D

Acq: 05 Jun 2023 19:41

Instrument :

MSVOA_V

ClientSampleId :

F6C29MS





Tgt Ion:	101	Resp:	6260
Ion Ratio	Lower	Upper	
101	100		
85	46.6	37.2	55.8
151	73.8	55.5	83.3

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 06/06/2023

Supervised By :Mahesh Dadoda 06/08/2023



#12

1,1-Dichloroethene

Concen: 4.303 ug/L




RT: 2.066 min Scan# 319


Delta R.T. -0.003 min

Lab File: VV031358.D

Acq: 05 Jun 2023 19:41

Tgt Ion:	96	Resp:	55829
Ion Ratio	Lower	Upper	
96	100		
61	194.5	138.3	256.8
63	126.7	101.9	189.3





#13

Acetone

Concen: 23.990 ug/L

RT: 2.224 min Scan# 30

Delta R.T. 0.061 min

Lab File: VV031358.D

Acq: 05 Jun 2023 19:41

Instrument :

MSVOA_V

ClientSampleId :

F6C29MS

Tgt Ion: 43 Resp: 9065

Ion Ratio Lower Upper

43 100





58 2.2 0.0 48.4

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 06/06/2023

Supervised By :Mahesh Dadoda 06/08/2023



#14

Carbon disulfide

Concen: 4.113 ug/L

RT: 2.237 min Scan# 372

Delta R.T. -0.003 min

Lab File: VV031358.D




Acq: 05 Jun 2023 19:41


Tgt Ion: 76 Resp: 158197

Ion Ratio Lower Upper

76 100


78 9.1 7.2 10.8





#15
Methyl Acetate
Concen: 3.547 ug/L m
RT: 2.413 min Scan# 431
Delta R.T. 0.022 min
Lab File: VV031358.D
Acq: 05 Jun 2023 19:41




Instrument :
MSVOA_V
ClientSampleId :
F6C29MS



Tgt Ion: 43 Resp: 2603
Ion Ratio Lower Upper
43 100
74 23.6 16.2 24.2




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/06/2023
Supervised By :Mahesh Dadoda 06/08/2023



#16
Methylene chloride
Concen: 4.225 ug/L
RT: 2.446 min Scan# 437
Delta R.T. -0.003 min
Lab File: VV031358.D
Acq: 05 Jun 2023 19:41

Tgt Ion: 84 Resp: 66642
Ion Ratio Lower Upper
84 100
86 67.1 46.7 86.7
49 141.7 101.6 188.8









#17
Methyl tert-butyl Ether
Concen: 4.744 ug/L
RT: 2.716 min Scan# 519
Delta R.T. 0.006 min
Lab File: VV031358.D
Acq: 05 Jun 2023 19:41

Instrument :
MSVOA_V
ClientSampleId :
F6C29MS

Tgt Ion: 73 Resp: 14027
Ion Ratio Lower Upper
73 100
43 30.6 17.8 26.8
57 62.4 19.8 29.8




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/06/2023
Supervised By :Mahesh Dadoda 06/08/2023



#18
trans-1,2-Dichloroethene
Concen: 4.660 ug/L
RT: 2.693 min Scan# 514
Delta R.T. -0.003 min
Lab File: VV031358.D
Acq: 05 Jun 2023 19:41


Tgt Ion: 96 Resp: 57342
Ion Ratio Lower Upper
96 100
61 155.5 114.2 212.0
98 65.0 45.0 83.6





#19
1,1-Dichloroethane
Concen: 4.263 ug/L
RT: 3.114 min Scan# 645
Delta R.T. 0.000 min
Lab File: VV031358.D
Acq: 05 Jun 2023 19:41




Instrument :
MSVOA_V
ClientSampleId :
F6C29MS



Tgt Ion: 63 Resp: 11651
Ion Ratio Lower Upper
63 100
65 31.5 22.0 40.8
83 12.9 9.0 16.6




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/06/2023
Supervised By :Mahesh Dadoda 06/08/2023



#21
2-Butanone
Concen: 34.960 ug/L m
RT: 3.944 min Scan# 903
Delta R.T. 0.035 min
Lab File: VV031358.D
Acq: 05 Jun 2023 19:41


Tgt Ion: 43 Resp: 171807
Ion Ratio Lower Upper
43 100
72 6.3 7.8 23.5#





#22
 cis-1,2-Dichloroethene
 Concen: 21.680 ug/L
 RT: 3.818 min Scan# 864
 Delta R.T. 0.000 min
 Lab File: VV031358.D
 Acq: 05 Jun 2023 19:41




Instrument :
 MSVOA_V
 ClientSampleId :
 F6C29MS



Tgt Ion: 96 Resp: 27032
 Ion Ratio Lower Upper
 96 100
 61 143.3 100.8 187.2
 98 62.3 45.1 83.9




Manual Integrations
 APPROVED


Reviewed By :Krupa Patel 06/06/2023
 Supervised By :Mahesh Dadoda 06/08/2023



#23
 Bromochloromethane
 Concen: 4.457 ug/L
 RT: 4.153 min Scan# 968
 Delta R.T. -0.000 min
 Lab File: VV031358.D
 Acq: 05 Jun 2023 19:41

Tgt Ion:128 Resp: 26271
 Ion Ratio Lower Upper
 128 100
 49 194.7 148.1 275.0
 130 123.4 89.9 166.9
 51 58.7 50.6 76.0









#25
Chloroform
Concen: 4.066 ug/L
RT: 4.282 min Scan# 1008
Delta R.T. -0.000 min
Lab File: VV031358.D
Acq: 05 Jun 2023 19:41

Instrument :
MSVOA_V
ClientSampleId :
F6C29MS

Tgt Ion: 83 Resp: 116194
Ion Ratio Lower Upper
83 100
85 66.9 44.9 83.3




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/06/2023
Supervised By :Mahesh Dadoda 06/08/2023



#27
1,2-Dichloroethane
Concen: 5.024 ug/L
RT: 5.050 min Scan# 1247
Delta R.T. 0.003 min
Lab File: VV031358.D
Acq: 05 Jun 2023 19:41

Tgt Ion: 62 Resp: 83451
Ion Ratio Lower Upper
62 100
98 9.2 7.2 10.8





#29

1,1,1-Trichloroethane

Concen: 4.929 ug/L

RT: 4.516 min Scan# 1081

Delta R.T. -0.000 min

Lab File: VV031358.D

Acq: 05 Jun 2023 19:41

Instrument :

MSVOA_V

ClientSampleId :

F6C29MS

Tgt Ion: 97 Resp: 106198

Ion Ratio Lower Upper

97 100

99 64.1 52.2 78.4





61 46.1 36.4 54.6

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 06/06/2023

Supervised By :Mahesh Dadoda 06/08/2023



#30

Cyclohexane

Concen: 30.349 ug/L

RT: 4.584 min Scan# 1102

Delta R.T. -0.003 min

Lab File: VV031358.D

Acq: 05 Jun 2023 19:41




Tgt Ion: 56 Resp: 593677


Ion Ratio Lower Upper

56 100

69 28.6 23.2 34.8

84 77.4 58.5 87.7





#31

Carbon tetrachloride

Concen: 4.910 ug/L

RT: 4.738 min Scan# 1150

Delta R.T. -0.000 min

Lab File: VV031358.D

Acq: 05 Jun 2023 19:41

Instrument :

MSVOA_V

ClientSampleId :

F6C29MS

Tgt Ion:117 Resp: 94352

Ion Ratio Lower Upper

117 100





119 97.4 76.3 114.5

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 06/06/2023

Supervised By :Mahesh Dadoda 06/08/2023



#33

Benzene

Concen: 21.211 ug/L




RT: 5.011 min Scan# 1235


Delta R.T. -0.003 min

Lab File: VV031358.D

Acq: 05 Jun 2023 19:41

Tgt Ion: 78 Resp: 978193





#34

Trichloroethene

Concen: 33.210 ug/L

RT: 5.835 min Scan# 1491

Delta R.T. 0.000 min

Lab File: VV031358.D


Acq: 05 Jun 2023 19:41

Instrument :

MSVOA_V

ClientSampleId :

F6C29MS



Tgt Ion: 95 Resp: 41364

Ion Ratio Lower Upper

95 100

97 64.1 46.1 85.5

132 91.7 64.5 119.9




130 95.2 63.1 117.1

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 06/06/2023

Supervised By :Mahesh Dadoda 06/08/2023



#35

Methylcyclohexane

Concen: 17.109 ug/L

RT: 6.053 min Scan# 1559

Delta R.T. -0.000 min

Lab File: VV031358.D

Acq: 05 Jun 2023 19:41




Tgt Ion: 83 Resp: 317236


Ion Ratio Lower Upper

83 100

55 90.7 74.6 112.0

98 48.0 36.5 54.7





#37

1,2-Dichloropropane

Concen: 4.663 ug/L

RT: 6.098 min Scan# 1573

Delta R.T. -0.000 min

Lab File: VV031358.D

Acq: 05 Jun 2023 19:41

Instrument :

MSVOA_V

ClientSampleId :

F6C29MS

Tgt Ion: 63 Resp: 63280

Ion Ratio Lower Upper

63 100





112 4.1 3.8 5.6

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 06/06/2023

Supervised By :Mahesh Dadoda 06/08/2023



#38

Bromodichloromethane

Concen: 4.692 ug/L

RT: 6.439 min Scan# 1679

Delta R.T. -0.000 min

Lab File: VV031358.D

Acq: 05 Jun 2023 19:41




Tgt Ion: 83 Resp: 81429


Ion Ratio Lower Upper

83 100

85 64.4 44.3 82.3

127 8.2 6.5 9.7





#39

cis-1,3-Dichloropropene

Concen: 4.916 ug/L

RT: 6.963 min Scan# 1841

Delta R.T. 0.003 min

Lab File: VV031358.D

Acq: 05 Jun 2023 19:41

Instrument :

MSVOA_V

ClientSampleId :

F6C29MS





Tgt Ion: 75 Resp: 8757
 Ion Ratio Lower Upper
 75 100
 77 28.8 21.3 39.5

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 06/06/2023

Supervised By :Mahesh Dadoda 06/08/2023



#40

4-Methyl-2-pentanone

Concen: 53.835 ug/L




RT: 7.169 min Scan# 1906


Delta R.T. 0.003 min

Lab File: VV031358.D

Acq: 05 Jun 2023 19:41


Tgt Ion: 43 Resp: 428712
 Ion Ratio Lower Upper
 43 100
 58 35.4 30.2 45.4
 100 12.3 9.0 13.6





#42
Toluene
Concen: 5.174 ug/L
RT: 7.320 min Scan# 1953
Delta R.T. 0.000 min
Lab File: VV031358.D
Acq: 05 Jun 2023 19:41




Instrument :
MSVOA_V
ClientSampleId :
F6C29MS



Tgt Ion: 91 Resp: 25319
Ion Ratio Lower Upper
91 100
92 57.1 39.2 72.8




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/06/2023
Supervised By :Mahesh Dadoda 06/08/2023



#44
trans-1,3-Dichloropropene
Concen: 4.979 ug/L
RT: 7.587 min Scan# 2036
Delta R.T. -0.000 min
Lab File: VV031358.D
Acq: 05 Jun 2023 19:41


Tgt Ion: 75 Resp: 76304
Ion Ratio Lower Upper
75 100
77 31.8 22.8 42.4





#45
1,1,2-Trichloroethane
Concen: 4.684 ug/L
RT: 7.777 min Scan# 2095
Delta R.T. 0.003 min
Lab File: VV031358.D
Acq: 05 Jun 2023 19:41




Instrument :
MSVOA_V
ClientSampleId :
F6C29MS




Tgt Ion: 97 Resp: 4315
Ion Ratio Lower Upper
97 100
99 62.6 43.1 80.1
83 88.6 63.1 117.1
85 56.8 41.1 76.3

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/06/2023
Supervised By :Mahesh Dadoda 06/08/2023




#47
Tetrachloroethene
Concen: 11.299 ug/L
RT: 7.908 min Scan# 2136
Delta R.T. 0.000 min
Lab File: VV031358.D
Acq: 05 Jun 2023 19:41




Tgt Ion:164 Resp: 109668
Ion Ratio Lower Upper
164 100
129 95.2 66.0 122.6
131 92.6 65.3 121.3
166 126.4 90.1 167.3





#48
2-Hexanone
Concen: 48.587 ug/L
RT: 8.088 min Scan# 2191
Delta R.T. 0.003 min
Lab File: VV031358.D
Acq: 05 Jun 2023 19:41




Instrument :
MSVOA_V
ClientSampleId :
F6C29MS



Tgt Ion: 43 Resp: 30728
Ion Ratio Lower Upper
43 100
58 50.6 41.1 61.7
57 18.3 14.4 21.6
100 9.9 7.8 11.6




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/06/2023
Supervised By :Mahesh Dadoda 06/08/2023



#49
Dibromochloromethane
Concen: 4.923 ug/L
RT: 8.182 min Scan# 2221
Delta R.T. -0.000 min
Lab File: VV031358.D
Acq: 05 Jun 2023 19:41

Tgt Ion:129 Resp: 50646
Ion Ratio Lower Upper
129 100
127 80.2 55.0 102.2





#50

1,2-Dibromoethane

Concen: 4.944 ug/L

RT: 8.291 min Scan# 21

Delta R.T. 0.003 min

Lab File: VV031358.D

Acq: 05 Jun 2023 19:41

Instrument :

MSVOA_V

ClientSampleId :

F6C29MS





Tgt Ion:	107	Resp:	4212
Ion	Ratio	Lower	Upper
107	100		
109	93.4	62.2	115.4
188	3.8	3.0	4.4

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 06/06/2023

Supervised By :Mahesh Dadoda 06/08/2023



#51

Chlorobenzene

Concen: 4.854 ug/L




RT: 8.822 min Scan# 2420


Delta R.T. 0.003 min

Lab File: VV031358.D

Acq: 05 Jun 2023 19:41

Tgt Ion:	112	Resp:	147954
Ion	Ratio	Lower	Upper
112	100		
114	31.0	22.7	42.1
77	62.8	52.2	78.2









#52
Ethylbenzene
Concen: 8.862 ug/L
RT: 8.953 min Scan# 2461
Delta R.T. 0.003 min
Lab File: VV031358.D
Acq: 05 Jun 2023 19:41

Instrument :
MSVOA_V
ClientSampleId :
F6C29MS

Tgt Ion: 91 Resp: 45360
Ion Ratio Lower Upper
91 100
106 29.4 20.0 37.2




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/06/2023
Supervised By :Mahesh Dadoda 06/08/2023



#53
m,p-Xylene
Concen: 17.756 ug/L
RT: 9.079 min Scan# 2500
Delta R.T. -0.000 min
Lab File: VV031358.D
Acq: 05 Jun 2023 19:41


Tgt Ion:106 Resp: 326982
Ion Ratio Lower Upper
106 100
91 217.8 154.3 286.5





#54
o-Xylene
Concen: 7.254 ug/L
RT: 9.484 min Scan# 2626
Delta R.T. 0.000 min
Lab File: VV031358.D
Acq: 05 Jun 2023 19:41




Instrument :
MSVOA_V
ClientSampleId :
F6C29MS



Tgt Ion:106 Resp: 125199
Ion Ratio Lower Upper
106 100
91 225.1 159.0 295.4




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/06/2023
Supervised By :Mahesh Dadoda 06/08/2023



#55
Styrene
Concen: 5.509 ug/L
RT: 9.503 min Scan# 2632
Delta R.T. 0.003 min
Lab File: VV031358.D
Acq: 05 Jun 2023 19:41


Tgt Ion:104 Resp: 169108
Ion Ratio Lower Upper
104 100
78 49.1 35.9 66.7





#57
1,1,2,2-Tetrachloroethane
Concen: 4.824 ug/L
RT: 10.185 min Scan# 2844
Delta R.T. -0.000 min
Lab File: VV031358.D
Acq: 05 Jun 2023 19:41




Instrument : MSVOA_V
ClientSampleId : F6C29MS



Tgt Ion: 83 Resp: 52970
Ion Ratio Lower Upper
83 100
85 64.4 45.0 83.6
131 10.2 7.3 10.9




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/06/2023
Supervised By :Mahesh Dadoda 06/08/2023



#59
Bromoform
Concen: 4.879 ug/L
RT: 9.674 min Scan# 2685
Delta R.T. 0.003 min
Lab File: VV031358.D
Acq: 05 Jun 2023 19:41

Tgt Ion:173 Resp: 29753
Ion Ratio Lower Upper
173 100
175 46.5 38.9 58.3
254 10.6 7.8 11.8









#60
Isopropylbenzene
Concen: 5.235 ug/L
RT: 9.873 min Scan# 21
Delta R.T. -0.000 min
Lab File: VV031358.D
Acq: 05 Jun 2023 19:41

Instrument :
MSVOA_V
ClientSampleId :
F6C29MS

Tgt Ion:105 Resp: 284614
Ion Ratio Lower Upper
105 100
120 25.6 19.9 29.9
77 17.8 14.2 21.2




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 06/06/2023
Supervised By :Mahesh Dadoda 06/08/2023



#61
1,2,3-Trichloropropane
Concen: 4.293 ug/L
RT: 10.217 min Scan# 2854
Delta R.T. -0.000 min
Lab File: VV031358.D
Acq: 05 Jun 2023 19:41


Tgt Ion: 75 Resp: 39693
Ion Ratio Lower Upper
75 100
77 32.5 27.3 40.9
110 38.6 29.4 44.0





#62
1,3,5-Trimethylbenzene
Concen: 6.123 ug/L
RT: 10.484 min Scan# 2936
Delta R.T. 0.003 min
Lab File: VV031358.D
Acq: 05 Jun 2023 19:41




Instrument :
MSVOA_V
ClientSampleId :
F6C29MS




Tgt Ion:105 Resp: 269740
Ion Ratio Lower Upper
105 100
120 46.1 35.9 53.9

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/06/2023
Supervised By :Mahesh Dadoda 06/08/2023




#63
1,2,4-Trimethylbenzene
Concen: 7.438 ug/L
RT: 10.860 min Scan# 3054
Delta R.T. 0.003 min
Lab File: VV031358.D
Acq: 05 Jun 2023 19:41




Tgt Ion:105 Resp: 338226
Ion Ratio Lower Upper
105 100
120 43.7 33.2 49.8





#64
1,3-Dichlorobenzene
Concen: 4.897 ug/L
RT: 11.124 min Scan# 3136
Delta R.T. -0.000 min
Lab File: VV031358.D
Acq: 05 Jun 2023 19:41




Instrument :
MSVOA_V
ClientSampleId :
F6C29MS




Tgt Ion:146 Resp: 128923
Ion Ratio Lower Upper
146 100
111 40.7 31.1 57.9
148 64.3 43.6 81.0

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/06/2023
Supervised By :Mahesh Dadoda 06/08/2023




#65
1,4-Dichlorobenzene
Concen: 4.689 ug/L
RT: 11.217 min Scan# 3165
Delta R.T. 0.003 min
Lab File: VV031358.D
Acq: 05 Jun 2023 19:41




Tgt Ion:146 Resp: 129308
Ion Ratio Lower Upper
146 100
111 39.3 28.4 52.8
148 63.9 44.2 82.0





#67
1,2-Dichlorobenzene
Concen: 4.772 ug/L
RT: 11.587 min Scan# 31
Delta R.T. -0.000 min
Lab File: VV031358.D
Acq: 05 Jun 2023 19:41




Instrument :
MSVOA_V
ClientSampleId :
F6C29MS




Tgt Ion:146 Resp: 120524
Ion Ratio Lower Upper
146 100
111 44.5 36.6 54.8
148 64.5 51.1 76.7

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/06/2023
Supervised By :Mahesh Dadoda 06/08/2023




#68
1,2-Dibromo-3-chloropropane
Concen: 4.357 ug/L
RT: 12.374 min Scan# 3525
Delta R.T. -0.000 min
Lab File: VV031358.D
Acq: 05 Jun 2023 19:41




Tgt Ion: 75 Resp: 9241
Ion Ratio Lower Upper
75 100
155 75.7 54.6 82.0
157 101.4 72.7 109.1





#69
1,3,5-Trichlorobenzene
Concen: 4.779 ug/L
RT: 12.590 min Scan# 3592
Delta R.T. -0.000 min
Lab File: VV031358.D
Acq: 05 Jun 2023 19:41




Instrument :
MSVOA_V
ClientSampleId :
F6C29MS




Tgt Ion:180 Resp: 93492
Ion Ratio Lower Upper
180 100
182 96.7 76.1 114.1
184 30.5 23.8 35.6
145 30.3 25.3 37.9

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 06/06/2023
Supervised By :Mahesh Dadoda 06/08/2023




#70
1,2,4-trichlorobenzene
Concen: 4.743 ug/L
RT: 13.207 min Scan# 3784
Delta R.T. 0.003 min
Lab File: VV031358.D
Acq: 05 Jun 2023 19:41




Tgt Ion:180 Resp: 74416
Ion Ratio Lower Upper
180 100
182 94.5 76.7 115.1
145 30.3 26.3 39.5





#71
Naphthalene
Concen: 5.127 ug/L
RT: 13.445 min Scan# 3859
Delta R.T. -0.003 min
Lab File: VV031358.D
Acq: 05 Jun 2023 19:41




Instrument :
MSVOA_V
ClientSampleId :
F6C29MS



Tgt Ion:128 Resp: 123412
Ion Ratio Lower Upper
128 100
129 11.1 9.0 13.6
127 12.8 11.4 17.2

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 06/06/2023
Supervised By :Mahesh Dadoda 06/08/2023



#72
1,2,3-Trichlorobenzene
Concen: 4.833 ug/L
RT: 13.689 min Scan# 3934
Delta R.T. 0.003 min
Lab File: VV031358.D
Acq: 05 Jun 2023 19:41

Tgt Ion:180 Resp: 68024
Ion Ratio Lower Upper
180 100
182 95.2 75.4 113.0
145 33.1 27.8 41.6

