

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV060619\
 Data File : VV011317.D
 Acq On : 08 Jun 2019 06:15
 Operator : SY/MD
 Sample : K3228-19
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F9P63

Quant Time: Jun 10 07:24:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Jun 10 06:50:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	160830	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	146483	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	61525	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	42821	3.51	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	70.20%
7) Chloroethane-d5	1.58	69	37053	3.66	ug/L	0.03
Spiked Amount	5.000	Range	65 - 130	Recovery	=	73.20%
11) 1,1-Dichloroethene-d2	2.12	63	69386	2.85	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	57.00%#
20) 2-Butanone-d5	3.98	46	128282	46.76	ug/L	0.04
Spiked Amount	50.000	Range	40 - 130	Recovery	=	93.52%
24) Chloroform-d	4.40	84	89366	4.17	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.40%
26) 1,2-Dichloroethane-d4	5.08	65	50755	4.38	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.60%
32) Benzene-d6	5.09	84	176265	4.02	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.40%
36) 1,2-Dichloropropane-d6	6.11	67	56554	4.19	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	83.80%
41) Toluene-d8	7.36	98	154175	3.83	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	76.60%
43) trans-1,3-Dichloropropene-	7.66	79	14086	2.74	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	54.80%#
46) 2-Hexanone-d5	8.14	63	102812	47.30	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.60%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	39749	4.32	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.40%
64) 1,2-Dichlorobenzene-d4	11.67	152	52172	4.30	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	86.00%

Target Compounds					Ovalue
13) Acetone	2.24	43	12155	6.608 ug/L	72
14) Carbon disulfide	2.31	76	102502	3.335 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV060619\
Data File : VV011317.D
Acq On : 08 Jun 2019 06:15
Operator : SY/MD
Sample : K3228-19
Misc : 25ML/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F9P63

Quant Time: Jun 10 07:24:53 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Jun 10 06:50:13 2019
Response via : Initial Calibration

