

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV060619\
 Data File : VV011291.D
 Acq On : 07 Jun 2019 16:42
 Operator : SY/MD
 Sample : VIBLK43
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK43

Quant Time: Jun 08 01:13:21 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Jun 08 00:40:24 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	191723	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	175259	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	70491	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	42849	2.95	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	59.00%
7) Chloroethane-d5	1.56	69	38163	3.16	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	63.20%#
11) 1,1-Dichloroethene-d2	2.12	63	69887	2.41	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	48.20%#
20) 2-Butanone-d5	3.94	46	123196	37.67	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	75.34%
24) Chloroform-d	4.40	84	87971	3.44	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	68.80%#
26) 1,2-Dichloroethane-d4	5.08	65	49064	3.55	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	71.00%
32) Benzene-d6	5.09	84	174829	3.33	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	66.60%#
36) 1,2-Dichloropropane-d6	6.12	67	52558	3.25	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	65.00%
41) Toluene-d8	7.36	98	152666	3.17	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	63.40%#
43) trans-1,3-Dichloropropene-	7.66	79	16465	2.67	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	53.40%#
46) 2-Hexanone-d5	8.13	63	93520	35.96	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	71.92%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	37400	3.40	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	68.00%
64) 1,2-Dichlorobenzene-d4	11.67	152	48799	3.51	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	70.20%#

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV060619\
Data File : VV011291.D
Acq On : 07 Jun 2019 16:42
Operator : SY/MD
Sample : VIBLK43
Misc : 25ML/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK43

Quant Time: Jun 08 01:13:21 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Jun 08 00:40:24 2019
Response via : Initial Calibration

