

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV060619\
 Data File : VV011310.D
 Acq On : 08 Jun 2019 02:59
 Operator : SY/MD
 Sample : VIBLK44
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK44

Quant Time: Jun 08 05:36:40 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Jun 08 05:28:44 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	148475	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	131217	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	51380	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	39368	3.50	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	70.00%
7) Chloroethane-d5	1.57	69	35656	3.81	ug/L	0.02
Spiked Amount	5.000	Range	65 - 130	Recovery	=	76.20%
11) 1,1-Dichloroethene-d2	2.12	63	61699	2.74	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	54.80%#
20) 2-Butanone-d5	3.95	46	141722	55.95	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	111.90%
24) Chloroform-d	4.40	84	83692	4.23	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.60%
26) 1,2-Dichloroethane-d4	5.08	65	51894	4.85	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.00%
32) Benzene-d6	5.09	84	164593	4.19	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.80%
36) 1,2-Dichloropropane-d6	6.12	67	53599	4.43	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	88.60%
41) Toluene-d8	7.36	98	137223	3.81	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	76.20%
43) trans-1,3-Dichloropropene-	7.66	79	15646	3.39	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	67.80%
46) 2-Hexanone-d5	8.14	63	106709	54.80	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	109.60%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	38635	4.69	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.80%
64) 1,2-Dichlorobenzene-d4	11.67	152	43428	4.28	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	85.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV060619\
Data File : VV011310.D
Acq On : 08 Jun 2019 02:59
Operator : SY/MD
Sample : VIBLK44
Misc : 25ML/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK44

Quant Time: Jun 08 05:36:40 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Jun 08 05:28:44 2019
Response via : Initial Calibration

