

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV060619\
 Data File : VV011318.D
 Acq On : 08 Jun 2019 06:43
 Operator : SY/MD
 Sample : VIBLK
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK

Quant Time: Jun 10 07:29:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Jun 10 06:50:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	177458	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	163195	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	65059	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	44118	3.28	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	65.60%
7) Chloroethane-d5	1.55	69	39600	3.54	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	70.80%
11) 1,1-Dichloroethene-d2	2.12	63	73691	2.74	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	54.80%#
20) 2-Butanone-d5	3.94	46	155777	51.46	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	102.92%
24) Chloroform-d	4.40	84	94010	3.97	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.40%
26) 1,2-Dichloroethane-d4	5.08	65	56934	4.45	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.00%
32) Benzene-d6	5.09	84	190574	3.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	78.00%
36) 1,2-Dichloropropane-d6	6.12	67	59293	3.94	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	78.80%
41) Toluene-d8	7.36	98	166240	3.71	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	74.20%
43) trans-1,3-Dichloropropene-	7.66	79	16712	2.92	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	58.40%
46) 2-Hexanone-d5	8.13	63	121105	50.01	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.02%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	45525	4.44	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.80%
64) 1,2-Dichlorobenzene-d4	11.67	152	54689	4.26	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	85.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV060619\
Data File : VV011318.D
Acq On : 08 Jun 2019 06:43
Operator : SY/MD
Sample : VIBLK
Misc : 25ML/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK

Quant Time: Jun 10 07:29:39 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Jun 10 06:50:13 2019
Response via : Initial Calibration

