

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV060620\
 Data File : VV016609.D
 Acq On : 07 Jun 2020 00:11
 Operator : SY/MD
 Sample : L2875-07
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ETTT0

Quant Time: Jun 07 01:54:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060320WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sun Jun 07 01:25:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	392961	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	359061	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	163101	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	122661	44.07	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	88.14%
7) Chloroethane-d5	1.57	69	82512	41.06	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	82.12%
11) 1,1-Dichloroethene-d2	2.12	63	160705	32.57	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	65.14%
21) 2-Butanone-d5	3.90	46	199110	111.88	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	111.88%
24) Chloroform-d	4.37	84	248855	47.64	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.28%
26) 1,2-Dichloroethane-d4	5.06	65	177046	50.62	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.24%
32) Benzene-d6	5.07	84	504731	52.48	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.96%
36) 1,2-Dichloropropane-d6	6.09	67	157986	52.72	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	105.44%
41) Toluene-d8	7.34	98	452218	51.77	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.54%
43) trans-1,3-Dichloropropene-	7.64	79	70808	48.48	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.96%
47) 2-Hexanone-d5	8.11	63	142953	111.38	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	111.38%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	199584	49.10	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	98.20%
64) 1,2-Dichlorobenzene-d4	11.65	152	169527	55.31	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	110.62%

Target Compounds					Ovalue
5) Vinyl chloride	1.32	62	5671	1.879 ug/L #	49
13) Acetone	2.18	43	1659	0.997 ug/L	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV060620\
Data File : VV016609.D
Acq On : 07 Jun 2020 00:11
Operator : SY/MD
Sample : L2875-07
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ETTT0

Quant Time: Jun 07 01:54:27 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060320WMA.M
Quant Title : VOC Analysis
QLast Update : Sun Jun 07 01:25:32 2020
Response via : Initial Calibration

