

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV060620\
 Data File : VV016616.D
 Acq On : 07 Jun 2020 02:56
 Operator : SY/MD
 Sample : L2875-17
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ETTT9

Quant Time: Jun 07 03:45:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060320WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sun Jun 07 03:25:55 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	376541	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	344365	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	158723	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	116739	43.77	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	87.54%
7) Chloroethane-d5	1.56	69	58358	30.31	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	60.62%#
11) 1,1-Dichloroethene-d2	2.12	63	156443	33.09	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	66.18%
21) 2-Butanone-d5	3.90	46	187498	109.95	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	109.95%
24) Chloroform-d	4.37	84	240193	47.99	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.98%
26) 1,2-Dichloroethane-d4	5.05	65	173376	51.73	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.46%
32) Benzene-d6	5.07	84	485841	52.67	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.34%
36) 1,2-Dichloropropane-d6	6.09	67	153190	53.30	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	106.60%
41) Toluene-d8	7.34	98	438405	52.33	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	104.66%
43) trans-1,3-Dichloropropene-	7.64	79	66438	47.43	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.86%
47) 2-Hexanone-d5	8.11	63	136472	110.87	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	110.87%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	193573	49.66	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	99.32%
64) 1,2-Dichlorobenzene-d4	11.65	152	167731	56.23	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	112.46%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV060620\
Data File : VV016616.D
Acq On : 07 Jun 2020 02:56
Operator : SY/MD
Sample : L2875-17
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ETTT9

Quant Time: Jun 07 03:45:31 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060320WMA.M
Quant Title : VOC Analysis
QLast Update : Sun Jun 07 03:25:55 2020
Response via : Initial Calibration

