

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV060622\
 Data File : VV025994.D
 Acq On : 06 Jun 2022 11:45
 Operator : SY/MD
 Sample : VV0606MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK220

Quant Time: Jun 07 02:26:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM051022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jun 07 02:25:55 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.613	114	297391	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	303381	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	137694	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.310	65	139150	36.255	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	72.500%
7) Chloroethane-d5	1.571	69	122928	36.471	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	72.940%
11) 1,1-Dichloroethene-d2	2.108	63	208337	31.036	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	62.080%
21) 2-Butanone-d5	3.892	46	113666	69.095	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	69.100%
24) Chloroform-d	4.346	84	255453	47.652	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.300%
26) 1,2-Dichloroethane-d4	5.027	65	155938	48.130	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.260%
32) Benzene-d6	5.047	84	489642	46.296	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.600%
36) 1,2-Dichloropropane-d6	6.066	67	164110	50.520	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.040%
41) Toluene-d8	7.313	98	429973	44.124	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.240%
43) trans-1,3-Dichloroprop...	7.619	79	67154	46.311	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.620%
47) 2-Hexanone-d5	8.088	63	110799	93.035	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	93.030%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	210301	48.546	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	97.100%
66) 1,2-Dichlorobenzene-d4	11.622	152	164750	54.545	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	109.100%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV060622\
Data File : VV025994.D
Acq On : 06 Jun 2022 11:45
Operator : SY/MD
Sample : VV0606MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK220

Quant Time: Jun 07 02:26:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM051022WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Jun 07 02:25:55 2022
Response via : Initial Calibration

