

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV060622\
 Data File : VV026000.D
 Acq On : 06 Jun 2022 14:28
 Operator : SY/MD
 Sample : N3173-03
 Misc : 5.00mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Jun 07 02:28:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM051022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jun 07 02:25:55 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	321174	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	325364	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	152447	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	154070	37.170	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	74.340%	
7) Chloroethane-d5	1.571	69	130262	35.785	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	71.580%	
11) 1,1-Dichloroethene-d2	2.108	63	229739	31.690	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	63.380%	
21) 2-Butanone-d5	3.905	46	138539	77.979	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	77.980%	
24) Chloroform-d	4.349	84	276938	47.834	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	95.660%	
26) 1,2-Dichloroethane-d4	5.030	65	167169	47.776	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	95.560%	
32) Benzene-d6	5.050	84	540478	47.650	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	95.300%	
36) 1,2-Dichloropropane-d6	6.069	67	180236	51.736	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	103.480%	
41) Toluene-d8	7.313	98	478701	45.805	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	91.620%	
43) trans-1,3-Dichloroprop...	7.619	79	70298	45.204	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	90.400%	
47) 2-Hexanone-d5	8.088	63	132431	103.685	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	103.690%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	234475	50.470	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	100.940%	
66) 1,2-Dichlorobenzene-d4	11.622	152	184014	55.028	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	110.060%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV060622\
Data File : VV026000.D
Acq On : 06 Jun 2022 14:28
Operator : SY/MD
Sample : N3173-03
Misc : 5.00mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Jun 07 02:28:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM051022WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Jun 07 02:25:55 2022
Response via : Initial Calibration

