

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW060623\
 Data File : VW031370.D
 Acq On : 06 Jun 2023 13:13
 Operator : SY/MD
 Sample : 03000-03DL 40X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F6C26DL

Quant Time: Jun 07 02:01:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR060123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jun 07 01:59:39 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	120867	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	125500	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.191	152	56279	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	49394	4.017	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	80.400%
7) Chloroethane-d5	1.532	69	46250	4.453	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.000%
11) 1,1-Dichloroethene-d2	2.060	65	27491	4.605	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	92.200%
20) 2-Butanone-d5	3.828	46	116804	41.046	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	82.100%
24) Chloroform-d	4.256	84	110203	5.331	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.600%
26) 1,2-Dichloroethane-d4	4.950	65	56273	5.076	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.600%
32) Benzene-d6	4.966	84	190466	4.992	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.800%
36) 1,2-Dichloropropane-d6	5.995	67	67784	5.260	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	105.200%
41) Toluene-d8	7.249	98	160591	4.482	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.600%
43) trans-1,3-Dichloroprop...	7.561	79	21058	4.488	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.800%
46) 2-Hexanone-d5	8.037	63	76014	40.332	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	80.660%
56) 1,1,2,2-Tetrachloroeth...	10.159	84	43506	4.441	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.800%
66) 1,2-Dichlorobenzene-d4	11.567	152	56193	5.294	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.800%
Target Compounds						
16) Methylene chloride	2.452	84	3395	0.247	ug/L	79
17) Methyl tert-butyl Ether	2.716	73	4355	0.169	ug/L	# 84
22) cis-1,2-Dichloroethene	3.828	96	9866	0.909	ug/L	97
34) Trichloroethene	5.841	95	30628	2.539	ug/L	94
47) Tetrachloroethene	7.908	164	124321	13.226	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV060623\
Data File : VV031370.D
Acq On : 06 Jun 2023 13:13
Operator : SY/MD
Sample : 03000-03DL 40X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F6C26DL

Quant Time: Jun 07 02:01:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR060123WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jun 07 01:59:39 2023
Response via : Initial Calibration

