

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VW060623\  
 Data File : VW031383.D  
 Acq On : 06 Jun 2023 18:16  
 Operator : SY/MD  
 Sample : 02990-13  
 Misc : 25.0mL/MSVOA\_V/WATER  
 ALS Vial : 24 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 VHBLK001

Quant Time: Jun 07 02:05:04 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR060123WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Wed Jun 07 01:59:39 2023  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	121868	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	129273	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.191	152	54756	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	48805	3.937	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	78.800%	
7) Chloroethane-d5	1.532	69	45446	4.340	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	86.800%	
11) 1,1-Dichloroethene-d2	2.060	65	25862	4.297	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	86.000%	
20) 2-Butanone-d5	3.822	46	124668	43.449	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	86.900%	
24) Chloroform-d	4.256	84	103342	4.958	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	99.200%	
26) 1,2-Dichloroethane-d4	4.950	65	54895	4.911	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	98.200%	
32) Benzene-d6	4.966	84	176086	4.481	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	89.600%	
36) 1,2-Dichloropropane-d6	5.995	67	62758	4.728	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	94.600%	
41) Toluene-d8	7.249	98	145762	3.950	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	79.000%	
43) trans-1,3-Dichloroprop...	7.561	79	20976	4.340	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	86.800%	
46) 2-Hexanone-d5	8.034	63	74113	38.176	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	76.360%	
56) 1,1,2,2-Tetrachloroeth...	10.159	84	40811	4.044	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	80.800%	
66) 1,2-Dichlorobenzene-d4	11.567	152	52186	5.053	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	101.000%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV060623\  
 Data File : VV031383.D  
 Acq On : 06 Jun 2023 18:16  
 Operator : SY/MD  
 Sample : 02990-13  
 Misc : 25.0mL/MSVOA\_V/WATER  
 ALS Vial : 24 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 VHBLK001

Quant Time: Jun 07 02:05:04 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR060123WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Wed Jun 07 01:59:39 2023  
 Response via : Initial Calibration

