

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV060721\
 Data File : VV021528.D
 Acq On : 07 Jun 2021 14:34
 Operator : SY/MD
 Sample : M2618-16
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Jun 08 03:19:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM052721WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jun 08 03:17:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	465070	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	438852	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	251502	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	154641	50.504	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	101.000%
7) Chloroethane-d5	1.568	69	100367	44.064	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	88.120%
11) 1,1-Dichloroethene-d2	2.108	63	197261	37.702	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.400%
21) 2-Butanone-d5	3.883	46	206398	94.002	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	94.000%
24) Chloroform-d	4.352	84	281000	46.101	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.200%
26) 1,2-Dichloroethane-d4	5.034	65	171050	48.721	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.440%
32) Benzene-d6	5.053	84	592495	48.261	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.520%
36) 1,2-Dichloropropane-d6	6.072	67	178255	47.016	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	94.040%
41) Toluene-d8	7.317	98	563645	48.189	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.380%
43) trans-1,3-Dichloroprop...	7.622	79	95611	47.123	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.240%
47) 2-Hexanone-d5	8.088	63	165662	95.158	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	95.160%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	248934	47.300	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	94.600%
66) 1,2-Dichlorobenzene-d4	11.625	152	250518	47.918	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.840%
Target Compounds						Qvalue
46) Tetrachloroethene	7.985	164	2776	0.884	ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV060721\
Data File : VV021528.D
Acq On : 07 Jun 2021 14:34
Operator : SY/MD
Sample : M2618-16
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Jun 08 03:19:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM052721WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Jun 08 03:17:01 2021
Response via : Initial Calibration

