

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW060723\
 Data File : VW031428.D
 Acq On : 08 Jun 2023 03:29
 Operator : SY/MD
 Sample : 03077-05
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0FB0

Quant Time: Jun 08 05:47:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM060723WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jun 08 05:16:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.539	114	245607	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	248536	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.191	152	116536	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	88509	40.959	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	81.920%
7) Chloroethane-d5	1.532	69	76769	43.880	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	87.760%
11) 1,1-Dichloroethene-d2	2.060	65	42486	41.484	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.960%
21) 2-Butanone-d5	3.799	46	139831	89.538	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	89.540%
24) Chloroform-d	4.256	84	188390	45.832	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.660%
26) 1,2-Dichloroethane-d4	4.947	65	128661	46.658	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.320%
32) Benzene-d6	4.966	84	328689	44.794	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.580%
36) 1,2-Dichloropropane-d6	5.995	67	119363	46.799	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	93.600%
41) Toluene-d8	7.249	98	285168	42.306	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	84.620%
43) trans-1,3-Dichloroprop...	7.558	79	52864	41.147	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.300%
47) 2-Hexanone-d5	8.034	63	72303	75.266	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	75.270%
56) 1,1,2,2-Tetrachloroeth...	10.159	84	139775	43.425	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	86.840%
66) 1,2-Dichlorobenzene-d4	11.567	152	109228	50.527	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.060%
Target Compounds						
					Qvalue	
13) Acetone	2.124	43	8589	6.148	ug/L	97
34) Trichloroethene	5.844	95	4210	2.437	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV060723\
Data File : VV031428.D
Acq On : 08 Jun 2023 03:29
Operator : SY/MD
Sample : 03077-05
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COFB0

Quant Time: Jun 08 05:47:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM060723WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jun 08 05:16:53 2023
Response via : Initial Calibration

