

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV060723\
 Data File : VV031394.D
 Acq On : 07 Jun 2023 13:17
 Operator : SY/MD
 Sample : VV0607WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK349

Quant Time: Jun 08 04:23:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM060723WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jun 08 04:14:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.539	114	284707	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	285511	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.191	152	139500	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	120324	48.034	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	96.060%	
7) Chloroethane-d5	1.532	69	98969	48.801	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	97.600%	
11) 1,1-Dichloroethene-d2	2.063	65	58216	49.036	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	98.080%	
21) 2-Butanone-d5	3.796	46	181742	100.393	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	100.390%	
24) Chloroform-d	4.256	84	218329	45.821	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	91.640%	
26) 1,2-Dichloroethane-d4	4.947	65	154797	48.427	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	96.860%	
32) Benzene-d6	4.966	84	420187	49.848	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	99.700%	
36) 1,2-Dichloropropane-d6	5.992	67	146920	50.143	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	100.280%	
41) Toluene-d8	7.249	98	365732	47.231	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	94.460%	
43) trans-1,3-Dichloroprop...	7.558	79	67583	45.791	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	91.580%	
47) 2-Hexanone-d5	8.031	63	99133	89.831	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	89.830%	
56) 1,1,2,2-Tetrachloroeth...	10.159	84	164679	44.536	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	89.080%	
66) 1,2-Dichlorobenzene-d4	11.567	152	135540	52.377	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	104.760%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV060723\
Data File : VV031394.D
Acq On : 07 Jun 2023 13:17
Operator : SY/MD
Sample : VV0607WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK349

Quant Time: Jun 08 04:23:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM060723WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jun 08 04:14:38 2023
Response via : Initial Calibration

