

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV060723\
 Data File : VV031396.D
 Acq On : 07 Jun 2023 14:05
 Operator : SY/MD
 Sample : 02961-16
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Jun 08 04:23:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM060723WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jun 08 04:14:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.539	114	251559	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	255106	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.191	152	123710	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	104668	47.290	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	94.580%	
7) Chloroethane-d5	1.532	69	90421	50.461	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	100.920%	
11) 1,1-Dichloroethene-d2	2.063	65	51081	48.696	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	97.400%	
21) 2-Butanone-d5	3.796	46	171209	107.036	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	107.040%	
24) Chloroform-d	4.256	84	210069	49.896	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	99.800%	
26) 1,2-Dichloroethane-d4	4.947	65	147887	52.362	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	104.720%	
32) Benzene-d6	4.966	84	382910	50.840	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	101.680%	
36) 1,2-Dichloropropane-d6	5.995	67	137087	52.364	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	104.720%	
41) Toluene-d8	7.249	98	335506	48.492	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	96.980%	
43) trans-1,3-Dichloroprop...	7.558	79	64691	49.055	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	98.120%	
47) 2-Hexanone-d5	8.034	63	91797	93.098	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	93.100%	
56) 1,1,2,2-Tetrachloroeth...	10.159	84	156274	47.300	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	94.600%	
66) 1,2-Dichlorobenzene-d4	11.567	152	127637	55.618	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	111.240%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV060723\
Data File : VV031396.D
Acq On : 07 Jun 2023 14:05
Operator : SY/MD
Sample : 02961-16
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Jun 08 04:23:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM060723WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jun 08 04:14:38 2023
Response via : Initial Calibration

