

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV060723\
 Data File : VV031401.D
 Acq On : 07 Jun 2023 16:03
 Operator : SY/MD
 Sample : 03076-03
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0F82

Quant Time: Jun 08 04:24:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM060723WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jun 08 04:14:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	170825	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	175303	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.191	152	84150	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	86455	57.522	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 115.040%		
7) Chloroethane-d5	1.532	69	75062	61.687	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 123.380%		
11) 1,1-Dichloroethene-d2	2.063	65	43165	60.598	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 121.200%		
21) 2-Butanone-d5	3.799	46	118745	109.322	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 109.320%		
24) Chloroform-d	4.256	84	172252	60.251	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 120.500%		
26) 1,2-Dichloroethane-d4	4.947	65	119677	62.400	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 124.800%		
32) Benzene-d6	4.966	84	308664	59.638	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 119.280%		
36) 1,2-Dichloropropane-d6	5.995	67	108386	60.247	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 120.500%#		
41) Toluene-d8	7.249	98	268374	56.447	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 112.900%		
43) trans-1,3-Dichloroprop...	7.558	79	51685	57.035	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 114.060%		
47) 2-Hexanone-d5	8.034	63	63985	94.432	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 94.430%		
56) 1,1,2,2-Tetrachloroeth...	10.159	84	128029	56.392	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 112.780%		
66) 1,2-Dichlorobenzene-d4	11.567	152	99736	63.892	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 127.780%#		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV060723\
Data File : VV031401.D
Acq On : 07 Jun 2023 16:03
Operator : SY/MD
Sample : 03076-03
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0F82

Quant Time: Jun 08 04:24:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM060723WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jun 08 04:14:38 2023
Response via : Initial Calibration

