

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV060819\
 Data File : VV011332.D
 Acq On : 08 Jun 2019 21:42
 Operator : SY/MD
 Sample : K3276-10
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DB8E6

Quant Time: Jun 09 00:44:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sun Jun 09 00:10:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	139938	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	127382	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	48587	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	42548	4.01	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	80.20%
7) Chloroethane-d5	1.56	69	37725	4.28	ug/L	0.01
Spiked Amount	5.000	Range	65 - 130	Recovery	=	85.60%
11) 1,1-Dichloroethene-d2	2.12	63	70768	3.34	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.80%
20) 2-Butanone-d5	3.95	46	114780	48.08	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	96.16%
24) Chloroform-d	4.40	84	85276	4.57	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.40%
26) 1,2-Dichloroethane-d4	5.08	65	47565	4.71	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.20%
32) Benzene-d6	5.09	84	174338	4.57	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.40%
36) 1,2-Dichloropropane-d6	6.11	67	53205	4.53	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	90.60%
41) Toluene-d8	7.36	98	152949	4.37	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.40%
43) trans-1,3-Dichloropropene-	7.66	79	15417	3.45	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	69.00%
46) 2-Hexanone-d5	8.14	63	87321	46.19	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.38%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	35379	4.42	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.40%
64) 1,2-Dichlorobenzene-d4	11.67	152	47131	4.91	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.20%

Target Compounds					Ovalue
13) Acetone	2.20	43	8110	5.067 ug/L	88
34) Trichloroethene	5.96	95	3765	0.360 ug/L	93
47) Tetrachloroethene	8.02	164	2553	0.337 ug/L	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV060819\
Data File : VV011332.D
Acq On : 08 Jun 2019 21:42
Operator : SY/MD
Sample : K3276-10
Misc : 25ML/MSVOA V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DB8E6

Quant Time: Jun 09 00:44:15 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sun Jun 09 00:10:13 2019
Response via : Initial Calibration

