

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV060820\
 Data File : VV016675.D
 Acq On : 08 Jun 2020 21:46
 Operator : SY/MD
 Sample : L2875-02DL 50X
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ETTS0DL

Quant Time: Jun 09 03:58:38 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jun 09 03:29:34 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	323211	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	302354	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	144918	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	105985	46.47	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	92.94%
7) Chloroethane-d5	1.56	69	74911	54.74	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	109.48%
11) 1,1-Dichloroethene-d2	2.12	63	141560	36.96	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.92%
21) 2-Butanone-d5	3.89	46	184135	109.09	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	109.09%
24) Chloroform-d	4.37	84	229316	49.45	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.90%
26) 1,2-Dichloroethane-d4	5.05	65	162877	51.64	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.28%
32) Benzene-d6	5.07	84	449259	51.55	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.10%
36) 1,2-Dichloropropane-d6	6.09	67	142544	52.47	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	104.94%
41) Toluene-d8	7.34	98	401024	50.14	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.28%
43) trans-1,3-Dichloropropene-	7.64	79	66257	46.83	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	93.66%
47) 2-Hexanone-d5	8.11	63	121116	102.25	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	102.25%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	189653	52.73	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	105.46%
64) 1,2-Dichlorobenzene-d4	11.65	152	168671	54.54	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	109.08%

Target Compounds

					Ovalue
5) Vinyl chloride	1.32	62	94227	40.889 ug/L	98
20) cis-1,2-Dichloroethene	3.94	96	189636	79.528 ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV060820\
Data File : VV016675.D
Acq On : 08 Jun 2020 21:46
Operator : SY/MD
Sample : L2875-02DL 50X
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ETTS0DL

Quant Time: Jun 09 03:58:38 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Jun 09 03:29:34 2020
Response via : Initial Calibration

