

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV060820\
 Data File : VV016685.D
 Acq On : 09 Jun 2020 01:44
 Operator : SY/MD
 Sample : L2895-01
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ETW41

Quant Time: Jun 09 04:49:25 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jun 09 03:29:34 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	336600	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	319049	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	144305	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	105684	44.50	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	89.00%
7) Chloroethane-d5	1.56	69	74428	52.22	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	104.44%
11) 1,1-Dichloroethene-d2	2.12	63	144609	36.26	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.52%
21) 2-Butanone-d5	3.90	46	181458	103.22	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	103.22%
24) Chloroform-d	4.37	84	230263	47.68	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.36%
26) 1,2-Dichloroethane-d4	5.05	65	165953	50.52	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.04%
32) Benzene-d6	5.07	84	457800	49.78	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.56%
36) 1,2-Dichloropropane-d6	6.09	67	143458	50.04	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.08%
41) Toluene-d8	7.34	98	407285	48.26	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.52%
43) trans-1,3-Dichloropropene-	7.64	79	66629	44.62	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.24%
47) 2-Hexanone-d5	8.11	63	120596	96.49	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	96.49%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	188400	49.64	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	99.28%
64) 1,2-Dichlorobenzene-d4	11.65	152	164605	53.45	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	106.90%

Target Compounds				Ovalue
25) Chloroform	4.40	83	15147	3.437 ug/L 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV060820\
Data File : VV016685.D
Acq On : 09 Jun 2020 01:44
Operator : SY/MD
Sample : L2895-01
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ETW41

Quant Time: Jun 09 04:49:25 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Jun 09 03:29:34 2020
Response via : Initial Calibration

