

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV060820\
 Data File : VV016689.D
 Acq On : 09 Jun 2020 03:18
 Operator : SY/MD
 Sample : L2895-13
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EW0LO

Quant Time: Jun 09 05:12:53 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jun 09 03:40:57 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	324583	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	300817	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	143474	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	91811	40.09	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	80.18%
7) Chloroethane-d5	1.56	69	68250	49.66	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	99.32%
11) 1,1-Dichloroethene-d2	2.12	63	130889	34.03	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	68.06%
21) 2-Butanone-d5	3.89	46	174306	102.83	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	102.83%
24) Chloroform-d	4.37	84	217415	46.68	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.36%
26) 1,2-Dichloroethane-d4	5.05	65	159417	50.33	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.66%
32) Benzene-d6	5.07	84	425364	49.05	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.10%
36) 1,2-Dichloropropane-d6	6.09	67	132150	48.89	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.78%
41) Toluene-d8	7.34	98	380089	47.77	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.54%
43) trans-1,3-Dichloropropene-	7.64	79	63159	44.86	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.72%
47) 2-Hexanone-d5	8.11	63	121296	102.93	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	102.93%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	183832	51.37	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	102.74%
64) 1,2-Dichlorobenzene-d4	11.65	152	160995	52.58	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.16%

Target Compounds

					Ovalue
13) Acetone	2.17	43	3303	2.957 ug/L	99
14) Carbon disulfide	2.31	76	6839	1.173 ug/L #	94
42) Toluene	7.41	91	8239	0.878 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV060820\
Data File : VV016689.D
Acq On : 09 Jun 2020 03:18
Operator : SY/MD
Sample : L2895-13
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EW0LO

Quant Time: Jun 09 05:12:53 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Jun 09 03:40:57 2020
Response via : Initial Calibration

