

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW060823\
 Data File : VW031456.D
 Acq On : 08 Jun 2023 18:13
 Operator : SY/MD
 Sample : 03077-16
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Jun 09 05:13:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM060723WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jun 09 05:07:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.539	114	239557	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	247902	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.191	152	115897	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	86716	41.142	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	82.280%		
7) Chloroethane-d5	1.532	69	84381	49.449	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	98.900%		
11) 1,1-Dichloroethene-d2	2.063	65	44621	44.669	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	89.340%		
21) 2-Butanone-d5	3.796	46	156106	102.484	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	102.480%		
24) Chloroform-d	4.256	84	206296	51.455	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	102.920%		
26) 1,2-Dichloroethane-d4	4.947	65	143543	53.370	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	106.740%		
32) Benzene-d6	4.966	84	360013	49.188	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	98.380%		
36) 1,2-Dichloropropane-d6	5.995	67	129649	50.962	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	101.920%		
41) Toluene-d8	7.249	98	305607	45.454	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	90.900%		
43) trans-1,3-Dichloroprop...	7.558	79	56102	43.779	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	87.560%		
47) 2-Hexanone-d5	8.034	63	80130	83.627	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	83.630%		
56) 1,1,2,2-Tetrachloroeth...	10.159	84	152705	47.563	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	95.120%		
66) 1,2-Dichlorobenzene-d4	11.567	152	126354	58.771	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	117.540%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV060823\
Data File : VV031456.D
Acq On : 08 Jun 2023 18:13
Operator : SY/MD
Sample : 03077-16
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Jun 09 05:13:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM060723WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jun 09 05:07:24 2023
Response via : Initial Calibration

