

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW060823\
 Data File : VW031482.D
 Acq On : 09 Jun 2023 05:13
 Operator : SY/MD
 Sample : BLK
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK36

Quant Time: Jun 09 16:57:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM060723WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jun 09 06:27:10 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	85709	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.793	117	74098	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.191	152	28216	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	42380	56.200	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	112.400%
7) Chloroethane-d5	1.532	69	46863	76.759	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	153.520%#
11) 1,1-Dichloroethene-d2	2.063	65	16842	47.124	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.240%
21) 2-Butanone-d5	3.796	46	146081	268.048	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	268.050%#
24) Chloroform-d	4.259	84	114855	80.070	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	160.140%#
26) 1,2-Dichloroethane-d4	4.950	65	101752	105.740	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	211.480%#
32) Benzene-d6	4.970	84	158671	72.530	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	145.060%#
36) 1,2-Dichloropropane-d6	5.995	67	74035	97.361	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	194.720%#
41) Toluene-d8	7.249	98	87655	43.617	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	87.240%
43) trans-1,3-Dichloroprop...	7.561	79	29816	77.841	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	155.680%#
47) 2-Hexanone-d5	8.034	63	56753	198.159	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	198.160%#
56) 1,1,2,2-Tetrachloroeth...	10.159	84	91364	95.206	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	190.420%#
66) 1,2-Dichlorobenzene-d4	0.000	152	0d	0.000	ug/L	
Spiked Amount	50.000	Range	80 - 120	Recovery	=	0.000%#
Target Compounds					Qvalue	
13) Acetone	2.121	43	3830	7.856	ug/L	86
16) Methylene chloride	2.455	84	2696	5.223	ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV060823\
Data File : VV031482.D
Acq On : 09 Jun 2023 05:13
Operator : SY/MD
Sample : BLK
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK36

Quant Time: Jun 09 16:57:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM060723WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jun 09 06:27:10 2023
Response via : Initial Calibration

