

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW060823\
 Data File : VW031462.D
 Acq On : 08 Jun 2023 21:17
 Operator : SY/MD
 Sample : 03106-02
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0FD0

Quant Time: Jun 09 06:34:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM060723WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jun 09 06:27:10 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	5.538	114	234435	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	245660	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.191	152	113849	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	79531	38.558	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	77.120%
7) Chloroethane-d5	1.532	69	77581	46.458	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	92.920%
11) 1,1-Dichloroethene-d2	2.063	65	40646	41.579	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	83.160%
21) 2-Butanone-d5	3.796	46	136214	91.379	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	91.380%
24) Chloroform-d	4.256	84	194137	49.480	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.960%
26) 1,2-Dichloroethane-d4	4.947	65	133852	50.854	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.700%
32) Benzene-d6	4.966	84	332231	45.807	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.620%
36) 1,2-Dichloropropane-d6	5.995	67	124410	49.349	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	98.700%
41) Toluene-d8	7.249	98	284127	42.645	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.280%
43) trans-1,3-Dichloroprop...	7.558	79	51315	40.409	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	80.820%
47) 2-Hexanone-d5	8.034	63	72852	76.725	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	76.730%
56) 1,1,2,2-Tetrachloroeth...	10.159	84	143045	44.961	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.920%
66) 1,2-Dichlorobenzene-d4	11.567	152	114278	54.110	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	108.220%
Target Compounds						Qvalue
34) Trichloroethene	5.841	95	20544	12.030	ug/L	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV060823\
Data File : VV031462.D
Acq On : 08 Jun 2023 21:17
Operator : SY/MD
Sample : 03106-02
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0FD0

Quant Time: Jun 09 06:34:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM060723WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jun 09 06:27:10 2023
Response via : Initial Calibration





#34
 Trichloroethene
 Concen: 12.030 ug/L
 RT: 5.841 min Scan# 1493
 Delta R.T. 0.003 min
 Lab File: VV031462.D
 Acq: 08 Jun 2023 21:17

Instrument :
 MSVOA_V
 ClientSampleId :
 C0FD0

Tgt Ion: 95	Resp: 20544
Ion Ratio	Lower Upper
95 100	
97 64.4	44.2 82.0
132 94.1	64.2 119.2
130 90.5	66.7 123.9

