

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV060920\
 Data File : VV016714.D
 Acq On : 09 Jun 2020 14:04
 Operator : SY/MD
 Sample : L2901-19
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ETTR2

Quant Time: Jun 10 03:36:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jun 10 00:34:30 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	340739	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	326729	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	156740	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	91038	37.86	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	75.72%
7) Chloroethane-d5	1.57	69	73338	50.83	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	101.66%
11) 1,1-Dichloroethene-d2	2.12	63	134358	33.28	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	66.56%
21) 2-Butanone-d5	3.89	46	185030	103.98	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	103.98%
24) Chloroform-d	4.37	84	232598	47.58	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.16%
26) 1,2-Dichloroethane-d4	5.05	65	168447	50.65	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.30%
32) Benzene-d6	5.07	84	451836	47.97	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.94%
36) 1,2-Dichloropropane-d6	6.09	67	143310	48.81	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.62%
41) Toluene-d8	7.34	98	400009	46.28	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.56%
43) trans-1,3-Dichloropropene-	7.64	79	66683	43.61	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.22%
47) 2-Hexanone-d5	8.11	63	119773	93.58	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	93.58%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	194805	50.12	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	100.24%
64) 1,2-Dichlorobenzene-d4	11.65	152	170432	50.95	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.90%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV060920\
Data File : VV016714.D
Acq On : 09 Jun 2020 14:04
Operator : SY/MD
Sample : L2901-19
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ETTR2

Quant Time: Jun 10 03:36:54 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jun 10 00:34:30 2020
Response via : Initial Calibration

